

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060722\
 Data File : VD073490.D
 Acq On : 07 Jun 2022 09:52
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleID :
 VSTDCCC050

Quant Time: Jun 07 16:32:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.500	0.505	-1.0	98	0.00
3 P	Chloromethane	0.565	0.477	15.6	82	0.00
4 C	Vinyl Chloride	0.517	0.468	9.5#	83	0.00
5 T	Bromomethane	0.374	0.341	8.8	86	0.00
6 T	Chloroethane	0.318	0.304	4.4	86	0.00
7 T	Trichlorofluoromethane	0.913	0.898	1.6	89	0.00
8 T	Diethyl Ether	0.251	0.242	3.6	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.537	1.5	90	0.00
10 T	Methyl Iodide	0.588	0.566	3.7	82	0.00
11 T	Tert butyl alcohol	0.086	0.060	30.2#	114	-0.01
12 CM	1,1-Dichloroethene	0.502	0.487	3.0#	85	0.00
13 T	Acrolein	0.039	0.018	53.8#	46#	0.00
14 T	Allyl chloride	0.811	0.732	9.7	79	0.00
15 T	Acrylonitrile	0.115	0.113	1.7	86	0.00
16 T	Acetone	0.096	0.097	-1.0	99	0.00
17 T	Carbon Disulfide	1.666	1.485	10.9	81	0.00
18 T	Methyl Acetate	0.266	0.249	6.4	85	0.00
19 T	Methyl tert-butyl Ether	1.145	1.175	-2.6	88	0.00
20 T	Methylene Chloride	0.716	0.585	18.3	84	0.00
21 T	trans-1,2-Dichloroethene	0.608	0.591	2.8	85	0.00
22 T	Diisopropyl ether	1.700	1.649	3.0	82	0.00
23 T	Vinyl Acetate	0.907	0.879	3.1	82	0.00
24 P	1,1-Dichloroethane	1.091	1.023	6.2	84	0.00
25 T	2-Butanone	0.140	0.135	3.6	87	0.00
26 T	2,2-Dichloropropane	0.931	0.913	1.9	88	0.00
27 T	cis-1,2-Dichloroethene	0.672	0.654	2.7	85	0.00
28 T	Bromochloromethane	0.376	0.320	14.9	74	0.00
29 T	Tetrahydrofuran	0.091	0.086	5.5	83	0.00
30 C	Chloroform	1.139	1.115	2.1#	87	0.00
31 T	Cyclohexane	0.966	0.843	12.7	79	0.00
32 T	1,1,1-Trichloroethane	0.997	0.979	1.8	88	0.00
33 S	1,2-Dichloroethane-d4	0.553	0.543	1.8	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.330	0.366	-10.9	90	0.00
36 T	1,1-Dichloropropene	0.496	0.511	-3.0	86	0.00
37 T	Ethyl Acetate	0.196	0.197	-0.5	86	0.00
38 T	Carbon Tetrachloride	0.503	0.550	-9.3	90	0.00
39 T	Methylcyclohexane	0.553	0.559	-1.1	81	0.00
40 TM	Benzene	1.430	1.501	-5.0	88	0.00
41 T	Methacrylonitrile	0.114	0.090	21.1	75	-0.01
42 TM	1,2-Dichloroethane	0.398	0.410	-3.0	86	0.00
43 T	Isopropyl Acetate	0.362	0.369	-1.9	84	0.00
44 TM	Trichloroethene	0.381	0.411	-7.9	90	0.00
45 C	1,2-Dichloropropane	0.360	0.368	-2.2#	85	0.00
46 T	Dibromomethane	0.198	0.210	-6.1	89	0.00
47 T	Bromodichloromethane	0.506	0.522	-3.2	87	0.00
48 T	Methyl methacrylate	0.182	0.175	3.8	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060722\
 Data File : VD073490.D
 Acq On : 07 Jun 2022 09:52
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 07 16:32:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	88	0.00
50 S	Toluene-d8	1.216	1.297	-6.7	85	0.00
51 T	4-Methyl-2-Pentanone	0.188	0.200	-6.4	86	0.00
52 CM	Toluene	0.904	0.993	-9.8#	88	0.00
53 T	t-1,3-Dichloropropene	0.457	0.484	-5.9	89	0.00
54 T	cis-1,3-Dichloropropene	0.547	0.573	-4.8	86	0.00
55 T	1,1,2-Trichloroethane	0.267	0.286	-7.1	90	0.00
56 T	Ethyl methacrylate	0.297	0.330	-11.1	85	0.00
57 T	1,3-Dichloropropane	0.449	0.475	-5.8	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.122	0.146	-19.7	91	0.00
59 T	2-Hexanone	0.124	0.142	-14.5	92	0.00
60 T	Dibromochloromethane	0.334	0.372	-11.4	93	0.00
61 T	1,2-Dibromoethane	0.256	0.278	-8.6	92	0.00
62 S	4-Bromofluorobenzene	0.427	0.494	-15.7	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.327	0.343	-4.9	92	0.00
65 PM	Chlorobenzene	1.026	1.064	-3.7	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.400	-5.3	91	0.00
67 C	Ethyl Benzene	1.801	1.893	-5.1#	88	0.00
68 T	m/p-Xylenes	0.700	0.755	-7.9	89	0.00
69 T	o-Xylene	0.643	0.683	-6.2	88	0.00
70 T	Styrene	1.106	1.227	-10.9	91	0.00
71 P	Bromoform	0.196	0.216	-10.2	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.564	3.505	1.7	88	0.00
74 T	N-amyl acetate	0.778	0.727	6.6	86	0.00
75 P	1,1,2,2-Tetrachloroethane	0.657	0.627	4.6	92	0.00
76 T	1,2,3-Trichloropropane	0.491	0.449	8.6	87	0.00
77 T	Bromobenzene	0.839	0.826	1.5	93	0.00
78 T	n-propylbenzene	4.475	4.423	1.2	89	0.00
79 T	2-Chlorotoluene	2.584	2.516	2.6	90	0.00
80 T	1,3,5-Trimethylbenzene	3.019	3.055	-1.2	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.204	0.196	3.9	88	0.00
82 T	4-Chlorotoluene	2.708	2.697	0.4	92	0.00
83 T	tert-Butylbenzene	2.545	2.632	-3.4	94	0.00
84 T	1,2,4-Trimethylbenzene	2.983	3.058	-2.5	94	0.00
85 T	sec-Butylbenzene	3.886	3.909	-0.6	92	0.00
86 T	p-Isopropyltoluene	3.167	3.266	-3.1	93	0.00
87 T	1,3-Dichlorobenzene	1.691	1.690	0.1	95	0.00
88 T	1,4-Dichlorobenzene	1.707	1.685	1.3	96	0.00
89 T	n-Butylbenzene	3.018	3.036	-0.6	90	0.00
90 T	Hexachloroethane	0.655	0.648	1.1	94	0.00
91 T	1,2-Dichlorobenzene	1.463	1.469	-0.4	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.101	0.096	5.0	93	0.00
93 T	1,2,4-Trichlorobenzene	0.831	0.858	-3.2	94	0.00
94 T	Hexachlorobutadiene	0.468	0.464	0.9	97	0.00
95 T	Naphthalene	1.466	1.545	-5.4	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060722\
Data File : VD073490.D
Acq On : 07 Jun 2022 09:52
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jun 07 16:32:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.737	0.747	-1.4	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6