

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD060818\
 Data File : VD058047.D
 Acq On : 9 Jun 2018 6:51
 Operator : VA/AP
 Sample : J3353-02
 Misc : 6.49g/5ml/MSVOA D/SOIL
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EPE-109-1(55-57)

Quant Time: Jun 11 04:22:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 08:57:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	146984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.68	114	246777	50.00	ug/l	-0.02
63) Chlorobenzene-d5	10.68	117	253525	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.90	152	147689	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.01	65	183563	124.92	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	249.84%	
35) Dibromofluoromethane	5.54	113	143349	75.08	ug/l	-0.02
Spiked Amount				50.000		
			Recovery	=	150.16%	
50) Toluene-d8	8.70	98	285963	63.00	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	126.00%	
62) 4-Bromofluorobenzene	11.93	95	152293	76.91	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	153.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.62	43	19561	92.39	ug/l #	87
20) Methylene Chloride	3.08	84	6464	5.86	ug/l #	79
30) Chloroform	5.32	83	21358	7.17	ug/l	85
95) Naphthalene	14.51	128	23060	4.84	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD060818\
Data File : VD058047.D
Acq On : 9 Jun 2018 6:51
Operator : VA/AP
Sample : J3353-02
Misc : 6.49g/5ml/MSVOA D/SOIL
ALS Vial : 66 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EPE-109-1(55-57)

Quant Time: Jun 11 04:22:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 08:57:33 2018
Response via : Initial Calibration

