

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD060818\
 Data File : VD058056.D
 Acq On : 9 Jun 2018 11:05
 Operator : VA/AP
 Sample : J3377-01RE
 Misc : 5.48g/5ml/MSVOA D/SOIL
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-B-DB-SRE

Quant Time: Jun 11 04:56:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 08:57:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	564493	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	6.68	114	746662	50.00	ug/l	-0.02
63) Chlorobenzene-d5	10.68	117	566957	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.90	152	269799	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.00	65	347819	61.63	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	123.26%	
35) Dibromofluoromethane	5.54	113	346170	59.92	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	119.84%	
50) Toluene-d8	8.69	98	764099	55.63	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	111.26%	
62) 4-Bromofluorobenzene	11.92	95	313768	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
Target Compounds						
30) Chloroform	5.32	83	12344	1.08	ug/l	Qvalue # 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD060818\
Data File : VD058056.D
Acq On : 9 Jun 2018 11:05
Operator : VA/AP
Sample : J3377-01RE
Misc : 5.48g/5ml/MSVOA D/SOIL
ALS Vial : 75 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-B-DB-SRE

Quant Time: Jun 11 04:56:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 08:57:33 2018
Response via : Initial Calibration

