

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD060820\
 Data File : VD065779.D
 Acq On : 08 Jun 2020 14:36
 Operator : VA/SY
 Sample : L2924-01
 Misc : 5.18G/5.00ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP

Quant Time: Jun 09 07:02:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060520S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 05 13:07:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.99	168	224817	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	389029	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	357476	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	164652	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	143856	64.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	129.52%	
35) Dibromofluoromethane	7.92	113	131987	53.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.42%	
50) Toluene-d8	10.34	98	483269	52.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.48%	
62) 4-Bromofluorobenzene	12.64	95	156119	50.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.48%	
Target Compounds						
16) Acetone	4.16	43	9438	22.523	ug/l #	89
20) Methylene Chloride	4.97	84	27524	8.692	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD060820\
Data File : VD065779.D
Acq On : 08 Jun 2020 14:36
Operator : VA/SY
Sample : L2924-01
Misc : 5.18G/5.00ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP

Quant Time: Jun 09 07:02:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060520S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 05 13:07:30 2020
Response via : Initial Calibration

