

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060821\
 Data File : VD069388.D
 Acq On : 08 Jun 2021 13:24
 Operator : VA/SY
 Sample : M2625-03RE
 Misc : 6.9G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 18-B12(146-150)RE

Quant Time: Jun 09 02:09:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 02 02:17:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	201119	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	369659	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	253699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	42947	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	144917	75.095	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	150.200%
35) Dibromofluoromethane	7.914	113	148892	64.140	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	128.280%
50) Toluene-d8	10.338	98	421188	49.896	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.800%
62) 4-Bromofluorobenzene	12.626	95	66965	24.210	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	48.420%
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	18049	53.596	ug/l	94
17) Carbon Disulfide	4.409	76	19654	2.946	ug/l	98
85) sec-Butylbenzene	13.396	105	3107	1.010	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060821\
Data File : VD069388.D
Acq On : 08 Jun 2021 13:24
Operator : VA/SY
Sample : M2625-03RE
Misc : 6.9G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
18-B12(146-150)RE

Quant Time: Jun 09 02:09:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060121S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 02 02:17:36 2021
Response via : Initial Calibration

