

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060922\
 Data File : VD073510.D
 Acq On : 09 Jun 2022 15:22
 Operator : VA/SY
 Sample : N3170-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20220413-COMPOSITE-B

Quant Time: Jun 09 17:35:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	199903	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	376093	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	385884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	150514	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	115669	53.919	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.840%	
35) Dibromofluoromethane	7.903	113	125723	51.985	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.980%	
50) Toluene-d8	10.332	98	452661	49.505	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.000%	
62) 4-Bromofluorobenzene	12.620	95	166799	50.249	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.500%	
Target Compounds						
						Qvalue
16) Acetone	4.156	43	23933	63.361	ug/l #	80
20) Methylene Chloride	4.967	84	13466	2.592	ug/l #	77
25) 2-Butanone	7.208	43	11315	20.359	ug/l	95
39) Methylcyclohexane	9.355	83	4376	1.034	ug/l #	82
86) p-Isopropyltoluene	13.502	119	12083	1.272	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD060922\
Data File : VD073510.D
Acq On : 09 Jun 2022 15:22
Operator : VA/SY
Sample : N3170-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20220413-COMPOSITE-B

Quant Time: Jun 09 17:35:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

