

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD061019\
 Data File : VD062656.D
 Acq On : 10 Jun 2019 16:27
 Operator : VA/AP
 Sample : K3273-17
 Misc : 3.85µl/5ml/MSVOA D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-9 0.5-1.0 060719

Manual Integrations
 APPROVED

MMDadoda
 6/14/2019 12:38:42 AM

Quant Time: Jun 11 08:04:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 00:02:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	263033	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.21	114	349827	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	223058	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	78648	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	215759	56.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.70%	
35) Dibromofluoromethane	6.92	113	185700	50.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.66%	
50) Toluene-d8	10.40	98	243744	31.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	62.82%	
62) 4-Bromofluorobenzene	13.55	95	89845	24.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	49.62%	
Target Compounds						
16) Acetone	3.22	43	25670	36.934	ug/l	95
67) Ethyl Benzene	12.52	91	55842	6.383	ug/l	100
68) m/p-Xylenes	12.66	106	28422	9.242	ug/l	96
69) o-Xylene	13.05	106	57423	20.331	ug/l	87
73) Isopropylbenzene	13.40	105	228985	41.307	ug/l	95
78) n-propylbenzene	13.77	91	501092	73.847	ug/l	97
80) 1,3,5-Trimethylbenzene	13.93	105	681446	139.091	ug/l	94
84) 1,2,4-Trimethylbenzene	14.23	105	869804	179.310	ug/l	97
85) sec-Butylbenzene	14.37	105	1025252	180.938	ug/l	99
86) p-Isopropyltoluene	14.49	119	626730	113.602	ug/l	99
89) n-Butylbenzene	14.79	91	483808m	97.480	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD061019\
Data File : VD062656.D
Acq On : 10 Jun 2019 16:27
Operator : VA/AP
Sample : K3273-17
Misc : 3.85g/5ml/MSVOA D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-9 0.5-1.0 060719

Manual Integrations
APPROVED

MMDadoda
6/14/2019 12:38:42 AM

Quant Time: Jun 11 08:04:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060719S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 00:02:21 2019
Response via : Initial Calibration

