

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD061219\
 Data File : VD062671.D
 Acq On : 12 Jun 2019 13:49
 Operator : VA/AP
 Sample : VD0612SBS02
 Misc : 4.36g/5ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBS01

Quant Time: Jun 13 07:23:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 00:02:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	775563	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	1030518	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	878914	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	485230	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.44	65	517445	46.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.48%	
35) Dibromofluoromethane	6.91	113	522653	48.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.18%	
50) Toluene-d8	10.39	98	1232189	53.89	ug/l	-0.02
Spiked Amount	50.000		Recovery	=	107.78%	
62) 4-Bromofluorobenzene	13.54	95	555679	52.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.20%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.56	85	142526	13.947	ug/l	90
3) Chloromethane	1.73	50	124525	16.227	ug/l	96
4) Vinyl Chloride	1.84	62	111556	14.098	ug/l	91
5) Bromomethane	2.13	94	53038	13.278	ug/l	94
6) Chloroethane	2.24	64	56224	14.391	ug/l #	79
7) Trichlorofluoromethane	2.51	101	241745	13.649	ug/l	100
8) Diethyl Ether	2.85	74	31085	12.904	ug/l	91
9) 1,1,2-Trichlorotrifluoroet	3.13	101	151462	15.857	ug/l	97
10) Methyl Iodide	3.29	142	157774	14.475	ug/l	92
11) Tert butyl alcohol	4.06	59	34367	74.708	ug/l #	84
12) 1,1-Dichloroethene	3.11	96	103406	15.085	ug/l	92
13) Acrolein	2.89	56	566	3.825	ug/l	88
14) Allyl chloride	3.60	41	161082	15.283	ug/l	93
15) Acrylonitrile	4.18	53	84204	80.551	ug/l	94
16) Acetone	3.21	43	150291	73.337	ug/l	96
17) Carbon Disulfide	3.35	76	286083	14.162	ug/l	98
18) Methyl Acetate	3.63	43	49459	14.686	ug/l	96
19) Methyl tert-butyl Ether	4.22	73	257538	15.404	ug/l	99
20) Methylene Chloride	3.80	84	102380	12.895	ug/l	95
21) trans-1,2-Dichloroethene	4.19	96	115045	15.300	ug/l	91
22) Diisopropyl ether	5.10	45	335821	15.345	ug/l	99
23) Vinyl Acetate	5.04	43	1023510	75.384	ug/l	100
24) 1,1-Dichloroethane	4.96	63	214231	15.216	ug/l	99
25) 2-Butanone	6.06	43	143501	71.777	ug/l	98
26) 2,2-Dichloropropane	6.00	77	246346	15.182	ug/l	96
27) cis-1,2-Dichloroethene	6.01	96	127595	15.653	ug/l	87
28) Bromochloromethane	6.42	49	84415	15.685	ug/l	97
29) Tetrahydrofuran	6.45	42	67429	74.893	ug/l	97
30) Chloroform	6.65	83	279846	15.492	ug/l	89
31) Cyclohexane	6.94	56	127688	14.458	ug/l	95
32) 1,1,1-Trichloroethane	6.85	97	285905	15.906	ug/l	98
36) 1,1-Dichloropropene	7.13	75	175358	14.868	ug/l	90
37) Ethyl Acetate	6.17	43	77544	16.250	ug/l #	87
38) Carbon Tetrachloride	7.10	117	293378	15.737	ug/l #	96

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD061219\
 Data File : VD062671.D
 Acq On : 12 Jun 2019 13:49
 Operator : VA/AP
 Sample : VD0612SBSD02
 Misc : 4.36g/5ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01

Quant Time: Jun 13 07:23:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 00:02:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	8.85	83	148153	15.545	ug/l	93
40) Benzene	7.44	78	359461	15.606	ug/l	98
41) Methacrylonitrile	6.43	41	90521	14.662	ug/l #	90
42) 1,2-Dichloroethane	7.56	62	198658	14.470	ug/l	94
43) Isopropyl Acetate	9.23	43	95568	15.357	ug/l #	84
44) Trichloroethene	8.52	130	145788	15.591	ug/l	94
45) 1,2-Dichloropropane	8.93	63	92142	16.655	ug/l	97
46) Dibromomethane	9.04	93	75734	15.422	ug/l	92
47) Bromodichloromethane	9.36	83	214193	16.034	ug/l	97
48) Methyl methacrylate	9.10	41	71705	15.949	ug/l	92
49) 1,4-Dioxane	9.06	88	12762	349.695	ug/l #	72
51) 4-Methyl-2-Pentanone	10.30	43	342387	78.629	ug/l	97
52) Toluene	10.49	92	226048	14.600	ug/l	98
53) t-1,3-Dichloropropene	10.91	75	176405	16.106	ug/l	99
54) cis-1,3-Dichloropropene	10.03	75	186824	16.503	ug/l	91
55) 1,1,2-Trichloroethane	11.18	97	81383	15.328	ug/l	93
56) Ethyl methacrylate	11.03	69	91545	17.122	ug/l	94
57) 1,3-Dichloropropane	11.40	76	132757	16.132	ug/l	92
58) 2-Chloroethyl Vinyl ether	9.84	63	313609	110.510	ug/l	100
59) 2-Hexanone	11.53	43	248929	76.851	ug/l	98
60) Dibromochloromethane	11.67	129	165523	15.792	ug/l	98
61) 1,2-Dibromoethane	11.79	107	99631	15.678	ug/l	99
64) Tetrachloroethene	11.25	164	125127	15.208	ug/l	91
65) Chlorobenzene	12.39	112	304239	15.649	ug/l	97
66) 1,1,1,2-Tetrachloroethane	12.51	131	144388	15.572	ug/l	98
67) Ethyl Benzene	12.52	91	547883	15.895	ug/l	98
68) m/p-Xylenes	12.66	106	375799	31.012	ug/l	95
69) o-Xylene	13.04	106	171237	15.386	ug/l	95
70) Styrene	13.06	104	314229	15.824	ug/l	95
71) Bromoform	13.23	173	103420	16.548	ug/l	97
73) Isopropylbenzene	13.40	105	537904	15.728	ug/l	100
74) N-amyl acetate	13.26	43	141890	15.351	ug/l	98
75) 1,1,2,2-Tetrachloroethane	13.69	83	92753	15.849	ug/l	99
76) 1,2,3-Trichloropropane	13.73	75	105830	16.085	ug/l	97
77) Bromobenzene	13.66	156	149548	15.298	ug/l	96
78) n-propylbenzene	13.77	91	690923	16.504	ug/l	99
79) 2-Chlorotoluene	13.83	91	407105	16.608	ug/l	98
80) 1,3,5-Trimethylbenzene	13.93	105	499982	16.541	ug/l	98
81) trans-1,4-Dichloro-2-buten	13.47	75	26772	16.547	ug/l	91
82) 4-Chlorotoluene	13.94	91	465612	15.884	ug/l	91
83) tert-Butylbenzene	14.18	119	559408	16.780	ug/l	96
84) 1,2,4-Trimethylbenzene	14.23	105	519489	17.358	ug/l	99
85) sec-Butylbenzene	14.36	105	510837	14.612	ug/l	97
86) p-Isopropyltoluene	14.49	119	568442	16.701	ug/l	95
87) 1,3-Dichlorobenzene	14.45	146	256636	14.492	ug/l	98
88) 1,4-Dichlorobenzene	14.53	146	281919	16.845	ug/l	98
89) n-Butylbenzene	14.79	91	494469	16.148	ug/l	97
90) Hexachloroethane	15.00	117	142613	15.871	ug/l	82
91) 1,2-Dichlorobenzene	14.80	146	240977	16.370	ug/l	94
92) 1,2-Dibromo-3-Chloropropan	15.37	75	20640	16.512	ug/l	72


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD061219\
Data File : VD062671.D
Acq On : 12 Jun 2019 13:49
Operator : VA/AP
Sample : VD0612SBSD02
Misc : 4.36g/5ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01

Quant Time: Jun 13 07:23:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060719S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 00:02:21 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	15.94	180	188210	17.564	ug/l	99
94) Hexachlorobutadiene	16.03	225	134627	15.490	ug/l	98
95) Naphthalene	16.11	128	249398	16.637	ug/l	97
96) 1,2,3-Trichlorobenzene	16.26	180	134260	15.834	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.04 min Scan# 570

Delta R.T. 0.00 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :


VD0612SBS01

Tot Ion:168 Resp: 775563

Ion Ratio Lower Upper

168 100

99 58.0 51.4 77.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC

7.04

250000

200000

150000



100000

50000

0

6.90 7.00 7.10 7.20

Time-->



#2

Dichlorodifluoromethane

Concen: 13.947 ug/l

RT: 1.56 min Scan# 14

Delta R.T. -0.01 min

Lab File: VD062671.D


Acq: 12 Jun 2019 13:49

Tgt Ion: 85 Resp: 142526

Ion Ratio Lower Upper

85 100

87 27.5 16.6 49.7



Abundance Ion 84.95 (84.65 to 85.65): VDC

Ion 86.90 (86.60 to 87.60): VDC

1.56

80000

60000


40000


20000

0

1.50 1.60 1.70

Time-->





#3

Chloromethane

Concen: 16.227 ug/l

RT: 1.73 min Scan# 31

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :


VD0612SBSD01

Tot Ion: 50 Resp: 124525

Ion Ratio Lower Upper




50 100

52 34.4 25.8 38.8



Abundance Ion 49.90 (49.60 to 50.60): VDC

Ion 51.90 (51.60 to 52.60): VDC



#4

Vinyl Chloride

Concen: 14.098 ug/l

RT: 1.84 min Scan# 42

Delta R.T. -0.01 min

Lab File: VD062671.D


Acq: 12 Jun 2019 13:49

Tgt Ion: 62 Resp: 111556

Ion Ratio Lower Upper



62 100


64 34.9 24.2 36.4



Abundance Ion 61.95 (61.65 to 62.65): VDC

Ion 63.95 (63.65 to 64.65): VDC





#5

Bromomethane

Concen: 13.278 ug/l

RT: 2.13 min Scan# 72

Delta R.T. -0.02 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion: 94 Resp: 53038

Ion Ratio Lower Upper




94 100

96 90.8 78.1 117.1

93 23.2 19.6 29.4



Abundance Ion 94.00 (93.70 to 94.70): VDC
Ion 96.00 (95.70 to 96.70): VDC
Ion 93.00 (92.70 to 93.70): VDC



#6

Chloroethane

Concen: 14.391 ug/l

RT: 2.24 min Scan# 83

Delta R.T. -0.02 min

Lab File: VD062671.D


Acq: 12 Jun 2019 13:49

Tgt Ion: 64 Resp: 56224



Ion Ratio Lower Upper


64 100

66 38.1 21.9 32.9#



Abundance Ion 63.95 (63.65 to 64.65): VDC
Ion 65.90 (65.60 to 66.60): VDC






#7




Trichlorofluoromethane
 Concen: 13.649 ug/l
 RT: 2.51 min Scan# 110
 Delta R.T. -0.02 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01

Tot Ion:101 Resp: 241745
 Ion Ratio Lower Upper
 101 100
 103 65.4 52.2 78.2




Abundance Ion 100.85 (100.55 to 101.55): V
 Ion 102.85 (102.55 to 103.55): V





#8


Diethyl Ether
 Concen: 12.904 ug/l
 RT: 2.85 min Scan# 145
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion: 74 Resp: 31085
 Ion Ratio Lower Upper
 74 100
 45 138.2 74.9 224.7



Abundance Ion 74.00 (73.70 to 74.70): VDC
 Ion 45.05 (44.75 to 45.75): VDC





#9

1,1,2-Trichlorotrifluoroethane

Concen: 15.857 ug/l

RT: 3.13 min Scan# 173

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion:101 Resp: 151462

Ion Ratio Lower Upper

101 100

85 43.3 36.0 54.0




151 68.3 52.6 79.0



Abundance Ion 100.85 (100.55 to 101.55): V

Ion 84.95 (84.65 to 85.65): V

Ion 150.90 (150.60 to 151.60): V



#10

Methyl Iodide

Concen: 14.475 ug/l

RT: 3.29 min Scan# 189

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49


Tgt Ion:142 Resp: 157774

Ion Ratio Lower Upper

142 100

127 62.2 55.2 82.8



141 17.0 12.1 18.1




Abundance Ion 141.85 (141.55 to 142.55): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.85 (140.55 to 141.55): V






#11




Tert butyl alcohol
Concen: 74.708 ug/l
RT: 4.06 min Scan# 268
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01

Tot Ion: 59 Resp: 34367
Ion Ratio Lower Upper
59 100
57 10.3 13.7 20.5#




Abundance Ion 58.95 (58.65 to 59.65): VDC
Ion 56.95 (56.65 to 57.65): VDC





#12


1,1-Dichloroethene
Concen: 15.085 ug/l
RT: 3.11 min Scan# 171
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 96 Resp: 103406
Ion Ratio Lower Upper
96 100
61 185.6 161.8 242.8
98 60.2 48.2 72.4



Abundance Ion 95.90 (95.60 to 96.60): VDC
Ion 60.95 (60.65 to 61.65): VDC
Ion 97.85 (97.55 to 98.55): VDC





#13

Acrolein

Concen: 3.825 ug/l

RT: 2.89 min Scan# 149

Delta R.T. -0.14 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :


VD0612SBS01

Tot Ion: 56 Resp: 566

Ion Ratio Lower Upper




56 100

55 84.3 59.4 89.0



Abundance Ion 55.95 (55.65 to 56.65): VDC

Ion 55.00 (54.70 to 55.70): VDC



#14

Allyl chloride

Concen: 15.283 ug/l

RT: 3.60 min Scan# 221

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49


Tgt Ion: 41 Resp: 161082

Ion Ratio Lower Upper

41 100

39 113.7 99.1 148.7



76 30.7 25.0 37.6




Abundance Ion 41.05 (40.75 to 41.75): VDC

Ion 38.95 (38.65 to 39.65): VDC

Ion 75.95 (75.65 to 76.65): VDC





#15

Acrylonitrile

Concen: 80.551 ug/l

RT: 4.18 min Scan# 280

Delta R.T. -0.02 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion: 53 Resp: 84204

Ion Ratio Lower Upper




53 100

52 93.5 73.6 110.4

51 42.5 42.0 63.0



Abundance Ion 53.00 (52.70 to 53.70): VDC
Ion 52.00 (51.70 to 52.70): VDC
Ion 51.00 (50.70 to 51.70): VDC



#16

Acetone

Concen: 73.337 ug/l

RT: 3.21 min Scan# 181

Delta R.T. -0.01 min

Lab File: VD062671.D


Acq: 12 Jun 2019 13:49

Tgt Ion: 43 Resp: 150291



Ion Ratio Lower Upper


43 100

58 15.0 10.8 16.2



Abundance Ion 42.95 (42.65 to 43.65): VDC
Ion 57.95 (57.65 to 58.65): VDC









#17
Carbon Disulfide
Concen: 14.162 ug/l
RT: 3.35 min Scan# 196
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01

Tot Ion: 76 Resp: 286083
Ion Ratio Lower Upper
76 100
78 9.5 7.0 10.4




Abundance Ion 75.85 (75.55 to 76.55): VDC
Ion 77.85 (77.55 to 78.55): VDC






#18
Methyl Acetate
Concen: 14.686 ug/l
RT: 3.63 min Scan# 224
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 43 Resp: 49459
Ion Ratio Lower Upper
43 100
74 16.8 12.1 18.1



Abundance Ion 42.95 (42.65 to 43.65): VDC
Ion 74.00 (73.70 to 74.70): VDC









#19
Methvl tert-butvl Ether
Concen: 15.404 ug/l
RT: 4.22 min Scan# 284
Delta R.T. -0.02 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01


Tot Ion: 73 Resp: 257538
Ion Ratio Lower Upper
73 100
57 16.4 13.5 20.3



#20
Methylene Chloride
Concen: 12.895 ug/l
RT: 3.80 min Scan# 241
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 84 Resp: 102380
Ion Ratio Lower Upper
84 100
49 133.6 111.2 166.8
51 41.3 31.9 47.9
86 61.3 44.7 67.1





#21

trans-1,2-Dichloroethene

Concen: 15.300 ug/l

RT: 4.19 min Scan# 281

Delta R.T. -0.02 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBS01


Tot Ion: 96 Resp: 115045

Ion Ratio Lower Upper

96 100

61 152.6 134.6 201.8




98 63.7 51.8 77.8



Abundance Ion 95.90 (95.60 to 96.60): VDC

Ion 60.95 (60.65 to 61.65): VDC

Ion 97.85 (97.55 to 98.55): VDC



#22

Diisopropyl ether

Concen: 15.345 ug/l

RT: 5.10 min Scan# 373

Delta R.T. -0.02 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Tgt Ion: 45 Resp: 335821


Ion Ratio Lower Upper

45 100

43 74.4 58.8 88.2

87 19.9 15.7 23.5

59 7.8 6.6 9.8






Abundance Ion 45.05 (44.75 to 45.75): VDC

Ion 43.05 (42.75 to 43.75): VDC

Ion 87.00 (86.70 to 87.70): VDC

Ion 58.95 (58.65 to 59.65): VDC





#23

Vinyl Acetate

Concen: 75.384 ug/l

RT: 5.04 min Scan# 367

Delta R.T. -0.02 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :


VD0612SBS01

Tot Ion: 43 Resp: 1023510

Ion Ratio Lower Upper




43 100

86 6.5 5.3 7.9



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 86.00 (85.70 to 86.70): VDC



#24

1,1-Dichloroethane

Concen: 15.216 ug/l

RT: 4.96 min Scan# 359

Delta R.T. -0.02 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49


Tgt Ion: 63 Resp: 214231

Ion Ratio Lower Upper

63 100

98 4.1 1.8 5.4



100 3.2 1.5 4.4




Abundance Ion 62.95 (62.65 to 63.65): VDC

Ion 97.95 (97.65 to 98.65): VDC

Ion 99.95 (99.65 to 100.65): VDC









#25
2-Butanone
Concen: 71.777 ug/l
RT: 6.06 min Scan# 471
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBS01


Tot Ion: 43 Resp: 143501
Ion Ratio Lower Upper
43 100
72 17.6 13.5 20.3



#26
2,2-Dichloropropane
Concen: 15.182 ug/l
RT: 6.00 min Scan# 465
Delta R.T. -0.02 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 77 Resp: 246346
Ion Ratio Lower Upper
77 100
97 16.1 8.8 26.5









#27
 cis-1,2-Dichloroethene
 Concen: 15.653 ug/l
 RT: 6.01 min Scan# 466
 Delta R.T. -0.02 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01

Tot Ion: 96 Resp: 127595
 Ion Ratio Lower Upper
 96 100
 61 151.4 0.0 340.0
 98 54.5 0.0 127.6



#28
 Bromochloromethane
 Concen: 15.685 ug/l
 RT: 6.42 min Scan# 507
 Delta R.T. -0.02 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion: 49 Resp: 84415
 Ion Ratio Lower Upper
 49 100
 129 0.0 0.0 8.2
 130 89.4 73.4 110.0





#29

Tetrahydrofuran

Concen: 74.893 ug/l

RT: 6.45 min Scan# 510

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion: 42 Resp: 67429

Ion Ratio Lower Upper




42 100

72 38.4 33.8 50.6

71 43.2 34.2 51.4



Abundance Ion 42.00 (41.70 to 42.70): VDC
Ion 72.00 (71.70 to 72.70): VDC
Ion 71.00 (70.70 to 71.70): VDC



#30

Chloroform

Concen: 15.492 ug/l

RT: 6.65 min Scan# 531

Delta R.T. -0.01 min

Lab File: VD062671.D


Acq: 12 Jun 2019 13:49

Tgt Ion: 83 Resp: 279846



Ion Ratio Lower Upper


83 100

85 60.8 55.8 83.6



Abundance Ion 82.95 (82.65 to 83.65): VDC
Ion 84.95 (84.65 to 85.65): VDC









#31
Cvclohexane
Concen: 14.458 ug/l
RT: 6.94 min Scan# 560
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01


Tot Ion: 56 Resp: 127688
Ion Ratio Lower Upper
56 100
69 33.3 26.8 40.2
84 86.3 73.9 110.9



#32
1,1,1-Trichloroethane
Concen: 15.906 ug/l
RT: 6.85 min Scan# 551
Delta R.T. -0.02 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 97 Resp: 285905
Ion Ratio Lower Upper
97 100
99 67.2 51.4 77.0
61 44.9 35.6 53.4









#33
 1,2-Dichloroethane-d4
 Concen: 46.236 ug/l
 RT: 7.44 min Scan# 611
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01


Tot Ion: 65 Resp: 517445
 Ion Ratio Lower Upper
 65 100
 67 42.6 0.0 81.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.21 min Scan# 689
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion: 114 Resp: 1030518
 Ion Ratio Lower Upper
 114 100
 63 18.8 0.0 42.4
 88 18.5 0.0 36.8





#35

Dibromofluoromethane

Concen: 48.090 ug/l

RT: 6.91 min Scan# 557

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion:113 Resp: 522653

Ion Ratio Lower Upper

113 100

111 101.1 76.4 114.6




192 16.4 11.4 17.0



Abundance Ion 112.90 (112.60 to 113.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 191.80 (191.50 to 192.50): V



#36

1,1-Dichloropropene

Concen: 14.868 ug/l

RT: 7.13 min Scan# 579

Delta R.T. -0.02 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49


Tgt Ion: 75 Resp: 175358

Ion Ratio Lower Upper

75 100

110 26.3 15.9 47.7



77 34.5 23.3 34.9




Abundance Ion 75.00 (74.70 to 75.70): V

Ion 109.90 (109.60 to 110.60): V

Ion 76.95 (76.65 to 77.65): V









#37
Ethyl Acetate
Concen: 16.250 ug/l
RT: 6.17 min Scan# 482
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01


Tot Ion: 43 Resp: 77544
Ion Ratio Lower Upper
43 100
61 18.5 8.7 13.1#
70 6.6 6.5 9.7



#38
Carbon Tetrachloride
Concen: 15.737 ug/l
RT: 7.10 min Scan# 576
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 117 Resp: 293378
Ion Ratio Lower Upper
117 100
119 90.6 73.5 110.3
121 24.0 24.2 36.4#









#39
Methvlcvclohexane
Concen: 15.545 ug/l
RT: 8.85 min Scan# 754
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01

Tot Ion: 83 Resp: 148153
Ion Ratio Lower Upper
83 100
55 100.9 84.6 127.0
98 56.8 38.2 57.4




Abundance Ion 83.05 (82.75 to 83.75): VDC
Ion 55.00 (54.70 to 55.70): VDC
Ion 98.05 (97.75 to 98.75): VDC





#40
Benzene
Concen: 15.606 ug/l
RT: 7.44 min Scan# 611
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 78 Resp: 359461
Ion Ratio Lower Upper
78 100
77 22.5 18.6 28.0



Abundance Ion 77.95 (77.65 to 78.65): VDC
Ion 76.95 (76.65 to 77.65): VDC





#41

Methacrylonitrile

Concen: 14.662 ug/l

RT: 6.43 min Scan# 508

Delta R.T. -0.02 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01

Tot Ion: 41 Resp: 90521


Ion Ratio Lower Upper

41 100

39 55.6 49.2 73.8

67 22.7 13.9 20.9#

52 16.1 7.6 11.4#






Abundance Ion 40.95 (40.65 to 41.65): VDC

Ion 38.95 (38.65 to 39.65): VDC

Ion 67.00 (66.70 to 67.70): VDC

Ion 52.00 (51.70 to 52.70): VDC



#42

1,2-Dichloroethane

Concen: 14.470 ug/l

RT: 7.56 min Scan# 623

Delta R.T. -0.02 min

Lab File: VD062671.D


Acq: 12 Jun 2019 13:49

Tgt Ion: 62 Resp: 198658

Ion Ratio Lower Upper



62 100


98 5.6 0.0 7.4



Abundance Ion 61.95 (61.65 to 62.65): VDC

Ion 97.85 (97.55 to 98.55): VDC





#43

Isopropyl Acetate

Concen: 15.357 ug/l

RT: 9.23 min Scan# 793

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01





Tot Ion: 43 Resp: 95568

Ion Ratio Lower Upper

43 100

61 31.9 19.0 28.6#

87 0.0 0.3 0.5#



#44

Trichloroethene

Concen: 15.591 ug/l

RT: 8.52 min Scan# 721

Delta R.T. -0.02 min

Lab File: VD062671.D




Acq: 12 Jun 2019 13:49

Tgt Ion: 130 Resp: 145788

Ion Ratio Lower Upper

130 100

95 98.3 0.0 185.6





#45

1,2-Dichloropropane

Concen: 16.655 ug/l

RT: 8.93 min Scan# 762

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :


VD0612SBSD01

Tot Ion: 63 Resp: 92142

Ion Ratio Lower Upper


63 100

65 29.4 24.6 37.0





Abundance Ion 62.95 (62.65 to 63.65): VDC

Ion 64.95 (64.65 to 65.65): VDC



Time--> 8.80 8.90 9.00



#46

Dibromomethane

Concen: 15.422 ug/l

RT: 9.04 min Scan# 774

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49


Tgt Ion: 93 Resp: 75734

Ion Ratio Lower Upper

93 100

95 79.1 68.7 103.1


174 107.6 79.8 119.8




Abundance Ion 92.90 (92.60 to 93.60): VDC


Ion 94.90 (94.60 to 95.60): VDC

Ion 173.80 (173.50 to 174.50): V



Time--> 8.90 9.00 9.10 9.20





#47

Bromodichloromethane

Concen: 16.034 ug/l

RT: 9.36 min Scan# 806

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion: 83 Resp: 214193

Ion Ratio Lower Upper




83 100

85 62.4 47.9 71.9

127 9.4 8.2 12.4



Abundance Ion 82.85 (82.55 to 83.55): VDC
Ion 84.95 (84.65 to 85.65): VDC
Ion 126.80 (126.50 to 127.50): VDC



#48

Methyl methacrylate

Concen: 15.949 ug/l

RT: 9.10 min Scan# 780

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49


Tgt Ion: 41 Resp: 71705

Ion Ratio Lower Upper



41 100


69 55.9 39.8 59.6

39 67.7 59.1 88.7



Abundance Ion 40.95 (40.65 to 41.65): VDC
Ion 69.00 (68.70 to 69.70): VDC
Ion 38.95 (38.65 to 39.65): VDC









#49
1,4-Dioxane
Concen: 349.695 ug/l
RT: 9.06 min Scan# 776
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01

Tot Ion: 88 Resp: 12762
Ion Ratio Lower Upper
88 100
43 62.7 40.7 61.1#
58 38.6 54.2 81.4#



#50
Toluene-d8
Concen: 53.895 ug/l
RT: 10.39 min Scan# 911
Delta R.T. -0.02 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 98 Resp: 1232189
Ion Ratio Lower Upper
98 100
100 66.7 52.1 78.1







#51
4-Methyl-2-Pentanone
Concen: 78.629 ug/l
RT: 10.30 min Scan# 901
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01



Tot Ion: 43 Resp: 342387
Ion Ratio Lower Upper
43 100
58 30.6 26.1 39.1



Abundance Ion 42.95 (42.65 to 43.65): VDC
Ion 57.95 (57.65 to 58.65): VDC




Time-->




#52
Toluene
Concen: 14.600 ug/l
RT: 10.49 min Scan# 921
Delta R.T. -0.02 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49


Tgt Ion: 92 Resp: 226048
Ion Ratio Lower Upper
92 100
91 174.2 137.0 205.6



Abundance Ion 92.00 (91.70 to 92.70): VDC
Ion 91.00 (90.70 to 91.70): VDC



Time-->









#53
 t-1,3-Dichloropropene
 Concen: 16.106 ug/l
 RT: 10.91 min Scan# 963
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01


Tot Ion: 75 Resp: 176405
 Ion Ratio Lower Upper
 75 100
 77 29.6 24.1 36.1



#54
 cis-1,3-Dichloropropene
 Concen: 16.503 ug/l
 RT: 10.03 min Scan# 874
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion: 75 Resp: 186824
 Ion Ratio Lower Upper
 75 100
 77 33.8 23.1 34.7
 39 72.0 63.4 95.2









#55
 1,1,2-Trichloroethane
 Concen: 15.328 ug/l
 RT: 11.18 min Scan# 991
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01

Tot Ion:	97	Resp:	81383
Ion	Ratio	Lower	Upper
97	100		
83	78.1	63.8	95.8
85	62.8	42.3	63.5
99	66.8	48.0	72.0




Abundance Ion 96.85 (96.55 to 97.55): VDC
 Ion 82.95 (82.65 to 83.65): VDC
 Ion 84.95 (84.65 to 85.65): VDC
 Ion 98.85 (98.55 to 99.55): VDC






#56
 Ethyl methacrylate
 Concen: 17.122 ug/l
 RT: 11.03 min Scan# 976
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion:	69	Resp:	91545
Ion	Ratio	Lower	Upper
69	100		
41	81.3	66.2	99.4
39	58.0	53.8	80.8



Abundance Ion 69.00 (68.70 to 69.70): VDC
 Ion 41.05 (40.75 to 41.75): VDC
 Ion 38.95 (38.65 to 39.65): VDC





#57

1,3-Dichloropropane

Concen: 16.132 ug/l

RT: 11.40 min Scan# 1013

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :





VD0612SBSD01

Tot Ion: 76 Resp: 132757

Ion Ratio Lower Upper

76 100

78 26.9 24.9 37.3



#58

2-Chloroethyl Vinyl ether

Concen: 110.510 ug/l

RT: 9.84 min Scan# 855

Delta R.T. -0.01 min

Lab File: VD062671.D




Acq: 12 Jun 2019 13:49


Tgt Ion: 63 Resp: 313609

Ion Ratio Lower Upper

63 100

106 25.4 20.5 30.7









#59
2-Hexanone
Concen: 76.851 ug/l
RT: 11.53 min Scan# 1026
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01


Tot Ion: 43 Resp: 248929
Ion Ratio Lower Upper
43 100
58 44.8 21.8 65.4



#60
Dibromochloromethane
Concen: 15.792 ug/l
RT: 11.67 min Scan# 1041
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 129 Resp: 165523
Ion Ratio Lower Upper
129 100
127 74.3 38.0 114.1









#61
 1,2-Dibromoethane
 Concen: 15.678 ug/l
 RT: 11.79 min Scan# 1053
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01

Tot Ion:107 Resp: 99631
 Ion Ratio Lower Upper
 107 100
 109 98.2 79.2 118.8




Abundance Ion 106.90 (106.60 to 107.60): V
 Ion 108.90 (108.60 to 109.60): V






#62
 4-Bromofluorobenzene
 Concen: 52.097 ug/l
 RT: 13.54 min Scan# 1231
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion: 95 Resp: 555679
 Ion Ratio Lower Upper
 95 100
 174 69.4 0.0 167.0
 176 69.4 0.0 159.2



Abundance Ion 95.00 (94.70 to 95.70): V
 Ion 173.90 (173.60 to 174.60): V
 Ion 175.90 (175.60 to 176.60): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 12.36 min Scan# 1111
Delta R.T. 0.00 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01

Tot Ion:117 Resp: 878914
Ion Ratio Lower Upper
117 100
82 52.0 48.7 73.1
119 32.7 27.0 40.6




Abundance Ion 116.95 (116.65 to 117.65): V
Ion 82.05 (81.75 to 82.75): V
Ion 118.95 (118.65 to 119.65): V






#64
Tetrachloroethene
Concen: 15.208 ug/l
RT: 11.25 min Scan# 998
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion:164 Resp: 125127
Ion Ratio Lower Upper
164 100
166 146.3 99.4 149.2
129 86.2 70.2 105.4
131 84.8 69.3 103.9



Abundance Ion 163.85 (163.55 to 164.55): V
Ion 165.85 (165.55 to 166.55): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V









#65
Chlorobenzene
Concen: 15.649 ug/l
RT: 12.39 min Scan# 1114
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01


Tot Ion:112 Resp: 304239
Ion Ratio Lower Upper
112 100
114 34.4 26.3 39.5



#66
1,1,1,2-Tetrachloroethane
Concen: 15.572 ug/l
RT: 12.51 min Scan# 1126
Delta R.T. 0.00 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion:131 Resp: 144388
Ion Ratio Lower Upper
131 100
133 91.9 46.8 140.3
119 61.6 29.9 89.8









#67
Ethyl Benzene
Concen: 15.895 ug/l
RT: 12.52 min Scan# 1127
Delta R.T. 0.00 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBS01


Tot Ion: 91 Resp: 547883
Ion Ratio Lower Upper
91 100
106 28.5 22.1 33.1



#68
m/p-Xylenes
Concen: 31.012 ug/l
RT: 12.66 min Scan# 1141
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion: 106 Resp: 375799
Ion Ratio Lower Upper
106 100
91 233.0 180.2 270.4









#69
o-Xylene
Concen: 15.386 ug/l
RT: 13.04 min Scan# 1180
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01


Tot Ion:106 Resp: 171237
Ion Ratio Lower Upper
106 100
91 239.0 124.1 372.1



#70
Styrene
Concen: 15.824 ug/l
RT: 13.06 min Scan# 1182
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion:104 Resp: 314229
Ion Ratio Lower Upper
104 100
78 54.6 45.9 68.9
103 43.4 38.4 57.6





#71

Bromoform

Concen: 16.548 ug/l

RT: 13.23 min Scan# 1199

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion:173 Resp: 103420

Ion Ratio Lower Upper

173 100

175 47.9 22.9 68.7




254 9.0 7.0 10.4



Abundance Ion 172.80 (172.50 to 173.50): V

Ion 174.80 (174.50 to 175.50): V

Ion 253.70 (253.40 to 254.40): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 14.51 min Scan# 1329

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49


Tgt Ion:152 Resp: 485230

Ion Ratio Lower Upper

152 100

115 62.2 30.1 90.5



150 160.5 0.0 314.0




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 115.00 (114.70 to 115.70): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 15.728 ug/l

RT: 13.40 min Scan# 1216

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :


VD0612SBSD01

Tot Ion:105 Resp: 537904

Ion Ratio Lower Upper




105 100

120 21.9 10.9 32.7



Abundance Ion 105.05 (104.75 to 105.75): V

Ion 120.05 (119.75 to 120.75): V



#74

N-ethyl acetate

Concen: 15.351 ug/l

RT: 13.26 min Scan# 1202

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Tgt Ion: 43 Resp: 141890


Ion Ratio Lower Upper

43 100

70 27.1 21.8 32.8

55 19.4 14.6 21.8

61 21.7 18.4 27.6






Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 70.00 (69.70 to 70.70): VDC

Ion 55.00 (54.70 to 55.70): VDC

Ion 60.95 (60.65 to 61.65): VDC





#75

1,1,2,2-Tetrachloroethane

Concen: 15.849 ug/l

RT: 13.69 min Scan# 1246

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion: 83 Resp: 92753

Ion Ratio Lower Upper

83 100

131 12.7 5.9 17.7




85 63.6 31.3 93.9



Abundance Ion 82.95 (82.65 to 83.65): VDC

Ion 130.90 (130.60 to 131.60): VDC

Ion 84.95 (84.65 to 85.65): VDC



#76

1,2,3-Trichloropropane

Concen: 16.085 ug/l

RT: 13.73 min Scan# 1250

Delta R.T. 0.00 min

Lab File: VD062671.D


Acq: 12 Jun 2019 13:49

Tgt Ion: 75 Resp: 105830

Ion Ratio Lower Upper



75 100

77 30.8 16.2 48.6



Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 76.95 (76.65 to 77.65): VDC





#77

Bromobenzene

Concen: 15.298 ug/l

RT: 13.66 min Scan# 1243

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion:156 Resp: 149548

Ion Ratio Lower Upper

156 100

77 117.6 56.3 168.8




158 102.7 49.6 148.8



Abundance Ion 155.90 (155.60 to 156.60): V

Ion 76.95 (76.65 to 77.65): V

Ion 157.85 (157.55 to 158.55): V



#78

n-propylbenzene

Concen: 16.504 ug/l

RT: 13.77 min Scan# 1254

Delta R.T. 0.00 min

Lab File: VD062671.D


Acq: 12 Jun 2019 13:49

Tgt Ion: 91 Resp: 690923

Ion Ratio Lower Upper



91 100


120 20.4 9.8 29.5



Abundance Ion 91.00 (90.70 to 91.70): V

Ion 120.05 (119.75 to 120.75): V









#79
 2-Chlorotoluene
 Concen: 16.608 ug/l
 RT: 13.83 min Scan# 1260
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01


Tot Ion: 91 Resp: 407105
 Ion Ratio Lower Upper
 91 100
 126 30.6 15.8 47.4



#80
 1,3,5-Trimethylbenzene
 Concen: 16.541 ug/l
 RT: 13.93 min Scan# 1270
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion: 105 Resp: 499982
 Ion Ratio Lower Upper
 105 100
 120 43.2 22.1 66.5





#81

trans-1,4-Dichloro-2-butene

Concen: 16.547 ug/l

RT: 13.47 min Scan# 1223

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01





Tot Ion: 75 Resp: 26772

Ion Ratio Lower Upper

75 100

53 145.8 109.4 164.2

89 55.4 37.0 55.6



#82

4-Chlorotoluene

Concen: 15.884 ug/l

RT: 13.94 min Scan# 1271

Delta R.T. -0.01 min

Lab File: VD062671.D




Acq: 12 Jun 2019 13:49


Tgt Ion: 91 Resp: 465612

Ion Ratio Lower Upper

91 100

126 26.5 15.6 46.8









#83
tert-Butylbenzene
Concen: 16.780 ug/l
RT: 14.18 min Scan# 1296
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49




Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01


Tot Ion:119 Resp: 559408
Ion Ratio Lower Upper
119 100
91 74.2 35.4 106.1
134 19.8 10.8 32.3



#84
1,2,4-Trimethylbenzene
Concen: 17.358 ug/l
RT: 14.23 min Scan# 1301
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion:105 Resp: 519489
Ion Ratio Lower Upper
105 100
120 39.7 20.3 60.8









#85
 sec-Butylbenzene
 Concen: 14.612 ug/l
 RT: 14.36 min Scan# 1314
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01

Tot Ion:105 Resp: 510837
 Ion Ratio Lower Upper
 105 100
 134 16.1 8.8 26.2




Abundance Ion 105.05 (104.75 to 105.75): V
 Ion 134.10 (133.80 to 134.80): V






#86
 p-Isopropyltoluene
 Concen: 16.701 ug/l
 RT: 14.49 min Scan# 1327
 Delta R.T. 0.00 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion:119 Resp: 568442
 Ion Ratio Lower Upper
 119 100
 134 21.0 10.9 32.6
 91 23.6 13.9 41.6



Abundance Ion 119.05 (118.75 to 119.75): V
 Ion 134.10 (133.80 to 134.80): V
 Ion 91.00 (90.70 to 91.70): V









#87
 1,3-Dichlorobenzene
 Concen: 14.492 ug/l
 RT: 14.45 min Scan# 1323
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01

Tot Ion	Ratio	Lower	Upper
146	100		
111	38.9	19.2	57.6
148	65.7	31.6	95.0




Abundance Ion 145.95 (145.65 to 146.65): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V






#88
 1,4-Dichlorobenzene
 Concen: 16.845 ug/l
 RT: 14.53 min Scan# 1331
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion	Ratio	Lower	Upper
146	100		
111	36.5	18.9	56.7
148	63.1	32.3	96.8



Abundance Ion 145.95 (145.65 to 146.65): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V





#89

n-Butylbenzene

Concen: 16.148 ug/l

RT: 14.79 min Scan# 1358

Delta R.T. 0.00 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion: 91 Resp: 494469

Ion Ratio Lower Upper

91 100

92 59.4 29.4 88.3




134 25.7 10.4 31.2



Abundance Ion 91.00 (90.70 to 91.70): V

Ion 92.00 (91.70 to 92.70): V

Ion 134.10 (133.80 to 134.80): V



#90

Hexachloroethane

Concen: 15.871 ug/l

RT: 15.00 min Scan# 1379

Delta R.T. -0.01 min

Lab File: VD062671.D


Acq: 12 Jun 2019 13:49

Tgt Ion: 117 Resp: 142613

Ion Ratio Lower Upper



117 100


201 66.7 41.5 124.6



Abundance Ion 116.85 (116.55 to 117.55): V

Ion 200.75 (200.45 to 201.45): V









#91
 1,2-Dichlorobenzene
 Concen: 16.370 ug/l
 RT: 14.80 min Scan# 1359
 Delta R.T. -0.01 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0612SBSD01


Tot Ion:146 Resp: 240977
 Ion Ratio Lower Upper
 146 100
 111 43.0 18.1 54.3
 148 64.4 31.4 94.0



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 16.512 ug/l
 RT: 15.37 min Scan# 1417
 Delta R.T. 0.00 min
 Lab File: VD062671.D
 Acq: 12 Jun 2019 13:49

Tgt Ion: 75 Resp: 20640
 Ion Ratio Lower Upper
 75 100
 155 72.5 30.1 90.5
 157 103.5 35.9 107.8









#93
1,2,4-Trichlorobenzene
Concen: 17.564 ug/l
RT: 15.94 min Scan# 1474
Delta R.T. -0.01 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Instrument :
MSVOA_D
ClientSampleId :
VD0612SBSD01

Tot Ion:180 Resp: 188210
Ion Ratio Lower Upper
180 100
182 98.0 49.4 148.2
145 29.1 14.5 43.6




Abundance Ion 179.85 (179.55 to 180.55): V
Ion 181.85 (181.55 to 182.55): V
Ion 144.95 (144.65 to 145.65): V






#94
Hexachlorobutadiene
Concen: 15.490 ug/l
RT: 16.03 min Scan# 1484
Delta R.T. 0.00 min
Lab File: VD062671.D
Acq: 12 Jun 2019 13:49

Tgt Ion:225 Resp: 134627
Ion Ratio Lower Upper
225 100
223 60.8 30.3 91.0
227 66.8 31.8 95.4



Abundance Ion 224.75 (224.45 to 225.45): V
Ion 222.75 (222.45 to 223.45): V
Ion 226.75 (226.45 to 227.45): V





#95

Naphthalene

Concen: 16.637 ug/l

RT: 16.11 min Scan# 1492

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49

Instrument :

MSVOA_D

ClientSampleId :

VD0612SBSD01


Tot Ion:128 Resp: 249398

Ion Ratio Lower Upper

128 100

127 11.2 9.7 14.5




129 11.6 8.1 12.1



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 15.834 ug/l

RT: 16.26 min Scan# 1507

Delta R.T. -0.01 min

Lab File: VD062671.D

Acq: 12 Jun 2019 13:49


Tgt Ion:180 Resp: 134260

Ion Ratio Lower Upper

180 100

182 104.2 47.7 143.1

145 29.3 15.0 45.0



Abundance Ion 179.85 (179.55 to 180.55): V

Ion 181.85 (181.55 to 182.55): V

Ion 144.95 (144.65 to 145.65): V

