

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061223\
 Data File : VD076395.D
 Acq On : 12 Jun 2023 15:48
 Operator : KP/SY
 Sample : VSTDICV050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD061223

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/13/2023
 Supervised By :Mahesh Dadoda 06/13/2023

Quant Time: Jun 13 02:06:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061223S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 13 02:01:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	292171	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	499251	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	442834	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	208313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	164891	48.861	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	97.720%		
35) Dibromofluoromethane	7.805	113	165741	51.365	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	102.740%		
50) Toluene-d8	10.269	98	613133	50.142	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	100.280%		
62) 4-Bromofluorobenzene	12.575	95	218821	52.163	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	104.320%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.928	85	139694	53.532	ug/l	97
3) Chloromethane	2.146	50	256536	51.917	ug/l	98
4) Vinyl Chloride	2.275	62	261930	55.398	ug/l	100
5) Bromomethane	2.681	94	135105	54.532	ug/l	100
6) Chloroethane	2.828	64	155889	54.750	ug/l	97
7) Trichlorofluoromethane	3.164	101	260529	54.757	ug/l	100
8) Diethyl Ether	3.587	74	88336	51.679	ug/l	69
9) 1,1,2-Trichlorotrifluo...	3.952	101	158326	55.344	ug/l	96
10) Methyl Iodide	4.152	142	195059	53.024	ug/l	91
11) Tert butyl alcohol	5.052	59	69524	246.443	ug/l #	93
12) 1,1-Dichloroethene	3.928	96	155481	53.293	ug/l #	77
13) Acrolein	3.793	56	72146	222.951	ug/l	98
14) Allyl chloride	4.552	41	290078	50.840	ug/l #	85
15) Acrylonitrile	5.252	53	219234	252.078	ug/l	99
16) Acetone	4.017	43	219052	295.272	ug/l #	77
17) Carbon Disulfide	4.264	76	448184	53.467	ug/l	98
18) Methyl Acetate	4.558	43	166414	49.201	ug/l #	83
19) Methyl tert-butyl Ether	5.311	73	423768	52.348	ug/l	96
20) Methylene Chloride	4.793	84	195986	50.726	ug/l #	73
21) trans-1,2-Dichloroethene	5.305	96	174930	51.596	ug/l #	80
22) Diisopropyl ether	6.216	45	599457	51.610	ug/l #	91
23) Vinyl Acetate	6.152	43	1276059m	255.796	ug/l	
24) 1,1-Dichloroethane	6.105	63	337835	52.327	ug/l	95
25) 2-Butanone	7.081	43	306542	264.192	ug/l #	74
26) 2,2-Dichloropropane	7.069	77	302847	52.098	ug/l	93
27) cis-1,2-Dichloroethene	7.075	96	209716	51.961	ug/l	83
28) Bromochloromethane	7.422	49	91344	51.276	ug/l #	69
29) Tetrahydrofuran	7.446	42	190835	257.795	ug/l #	75
30) Chloroform	7.593	83	333456	51.821	ug/l	97
31) Cyclohexane	7.875	56	282406	52.302	ug/l #	82
32) 1,1,1-Trichloroethane	7.787	97	297474	54.131	ug/l	98
36) 1,1-Dichloropropene	8.005	75	253110	53.742	ug/l	96
37) Ethyl Acetate	7.169	43	130699	52.760	ug/l #	91
38) Carbon Tetrachloride	7.987	117	244912	56.009	ug/l	98
39) Methylcyclohexane	9.269	83	303411	55.949	ug/l #	89
40) Benzene	8.252	78	721192	52.673	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061223\
 Data File : VD076395.D
 Acq On : 12 Jun 2023 15:48
 Operator : KP/SY
 Sample : VSTDICV050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD061223

Quant Time: Jun 13 02:06:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061223S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 13 02:01:34 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 06/13/2023
 Supervised By :Mahesh Dadoda 06/13/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.399	41	68874	43.461	ug/l #	80
42) 1,2-Dichloroethane	8.322	62	206225	50.579	ug/l	93
43) Isopropyl Acetate	8.357	43	270638	52.108	ug/l #	83
44) Trichloroethene	9.028	130	200542	54.030	ug/l	91
45) 1,2-Dichloropropane	9.305	63	193374	53.138	ug/l	99
46) Dibromomethane	9.387	93	103744	51.658	ug/l	95
47) Bromodichloromethane	9.581	83	261655	52.595	ug/l #	97
48) Methyl methacrylate	9.381	41	127898	51.378	ug/l #	69
49) 1,4-Dioxane	9.381	88	25775	1081.722	ug/l #	79
51) 4-Methyl-2-Pentanone	10.157	43	673394	261.041	ug/l #	83
52) Toluene	10.334	92	455491	52.624	ug/l	98
53) t-1,3-Dichloropropene	10.552	75	268654	52.798	ug/l	96
54) cis-1,3-Dichloropropene	10.016	75	306106	52.312	ug/l #	77
55) 1,1,2-Trichloroethane	10.734	97	140553	52.397	ug/l	97
56) Ethyl methacrylate	10.599	69	195432	52.461	ug/l #	68
57) 1,3-Dichloropropane	10.881	76	244617	52.421	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.869	63	367895	242.448	ug/l	93
59) 2-Hexanone	10.922	43	486994	269.321	ug/l	80
60) Dibromochloromethane	11.075	129	176851	53.284	ug/l	97
61) 1,2-Dibromoethane	11.181	107	135789	52.177	ug/l	98
64) Tetrachloroethene	10.810	164	163414	53.752	ug/l	93
65) Chlorobenzene	11.604	112	488732	53.244	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.681	131	181398	52.852	ug/l	96
67) Ethyl Benzene	11.681	91	909101	53.773	ug/l	96
68) m/p-Xylenes	11.793	106	685490	107.162	ug/l	87
69) o-Xylene	12.122	106	328679	53.488	ug/l	88
70) Styrene	12.134	104	557874	53.212	ug/l #	93
71) Bromoform	12.298	173	107148	53.523	ug/l #	98
73) Isopropylbenzene	12.422	105	861355	54.675	ug/l	96
74) N-amyl acetate	12.234	43	256923	53.477	ug/l #	82
75) 1,1,2,2-Tetrachloroethane	12.675	83	162473	52.226	ug/l	99
76) 1,2,3-Trichloropropane	12.728	75	92796m	47.649	ug/l	
77) Bromobenzene	12.704	156	199160	53.581	ug/l	84
78) n-propylbenzene	12.763	91	1055790	54.981	ug/l	96
79) 2-Chlorotoluene	12.851	91	582397	53.596	ug/l	94
80) 1,3,5-Trimethylbenzene	12.904	105	712001	54.147	ug/l	94
81) trans-1,4-Dichloro-2-b...	12.469	75	54384	49.999	ug/l	91
82) 4-Chlorotoluene	12.945	91	615102	54.054	ug/l	92
83) tert-Butylbenzene	13.169	119	614650	54.470	ug/l	93
84) 1,2,4-Trimethylbenzene	13.210	105	706517	54.381	ug/l	95
85) sec-Butylbenzene	13.345	105	918282	55.290	ug/l	97
86) p-Isopropyltoluene	13.463	119	778200	55.645	ug/l	95
87) 1,3-Dichlorobenzene	13.457	146	393401	54.688	ug/l	97
88) 1,4-Dichlorobenzene	13.540	146	386938	54.429	ug/l	98
89) n-Butylbenzene	13.787	91	753349	56.067	ug/l	98
90) Hexachloroethane	14.051	117	142068	55.048	ug/l	99
91) 1,2-Dichlorobenzene	13.834	146	340251	53.782	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.445	75	25471	49.841	ug/l	96
93) 1,2,4-Trichlorobenzene	15.098	180	232774	54.974	ug/l	98
94) Hexachlorobutadiene	15.204	225	119628	56.157	ug/l	99
95) Naphthalene	15.334	128	452446	53.212	ug/l	99
96) 1,2,3-Trichlorobenzene	15.522	180	201023	54.969	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061223\
Data File : VD076395.D
Acq On : 12 Jun 2023 15:48
Operator : KP/SY
Sample : VSTDICV050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023

Quant Time: Jun 13 02:06:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061223S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 13 02:01:34 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion:168 Resp: 29217
Ion Ratio Lower Upper
168 100
99 57.1 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#2
Dichlorodifluoromethane
Concen: 53.532 ug/l
RT: 1.928 min Scan# 58
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 85 Resp: 139694
Ion Ratio Lower Upper
85 100
87 32.8 15.7 47.0





#3
Chloromethane
Concen: 51.917 ug/l
RT: 2.146 min Scan# 95
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 50 Resp: 256530
Ion Ratio Lower Upper
50 100
52 32.2 26.6 40.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#4
Vinyl Chloride
Concen: 55.398 ug/l
RT: 2.275 min Scan# 117
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 62 Resp: 261930
Ion Ratio Lower Upper
62 100
64 32.8 26.0 39.0





#5
Bromomethane
Concen: 54.532 ug/l
RT: 2.681 min Scan# 186
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 94 Resp: 13510
Ion Ratio Lower Upper
94 100
96 95.2 75.8 113.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#6
Chloroethane
Concen: 54.750 ug/l
RT: 2.828 min Scan# 211
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 64 Resp: 155889
Ion Ratio Lower Upper
64 100
66 30.4 25.5 38.3





#7
 Trichlorofluoromethane
 Concen: 54.757 ug/l
 RT: 3.164 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VD076395.D
 Acq: 12 Jun 2023 15:48




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD061223



Tgt Ion:101 Resp: 260529
 Ion Ratio Lower Upper
 101 100
 103 65.2 52.3 78.5




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 06/13/2023
 Supervised By :Mahesh Dadoda 06/13/2023



#8
 Diethyl Ether
 Concen: 51.679 ug/l
 RT: 3.587 min Scan# 340
 Delta R.T. -0.006 min
 Lab File: VD076395.D
 Acq: 12 Jun 2023 15:48

Tgt Ion: 74 Resp: 88336
 Ion Ratio Lower Upper
 74 100
 45 101.8 37.5 112.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 55.344 ug/l
RT: 3.952 min Scan# 403
Delta R.T. -0.006 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48

Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223

Tgt Ion:101 Resp: 158320
Ion Ratio Lower Upper
101 100
85 45.5 33.8 50.6
151 76.9 59.8 89.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#10
Methyl Iodide
Concen: 53.024 ug/l
RT: 4.152 min Scan# 436
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion:142 Resp: 195059
Ion Ratio Lower Upper
142 100
127 43.9 29.4 44.0
141 14.3 11.1 16.7





#11
Tert butyl alcohol
Concen: 246.443 ug/l
RT: 5.052 min Scan# 589
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 59 Resp: 69524
Ion Ratio Lower Upper
59 100
57 12.4 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#12
1,1-Dichloroethene
Concen: 53.293 ug/l
RT: 3.928 min Scan# 398
Delta R.T. -0.006 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 96 Resp: 155481
Ion Ratio Lower Upper
96 100
61 162.0 100.6 150.8#
98 67.3 50.2 75.2





#13
Acrolein
Concen: 222.951 ug/l
RT: 3.793 min Scan# 31
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 56 Resp: 72140
Ion Ratio Lower Upper
56 100
55 70.5 55.1 82.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#14
Allyl chloride
Concen: 50.840 ug/l
RT: 4.552 min Scan# 504
Delta R.T. -0.006 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 41 Resp: 290078
Ion Ratio Lower Upper
41 100
39 69.4 53.8 80.6
76 35.2 45.8 68.8#





#15
Acrylonitrile
Concen: 252.078 ug/l
RT: 5.252 min Scan# 61
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion: 53 Resp: 219234
Ion Ratio Lower Upper
53 100
52 81.7 65.0 97.4
51 37.1 28.3 42.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#16
Acetone
Concen: 295.272 ug/l
RT: 4.017 min Scan# 413
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 43 Resp: 219052
Ion Ratio Lower Upper
43 100
58 28.8 34.7 52.1#





#17
Carbon Disulfide
Concen: 53.467 ug/l
RT: 4.264 min Scan# 41
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion: 76 Resp: 448184
Ion Ratio Lower Upper
76 100
78 9.0 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#18
Methyl Acetate
Concen: 49.201 ug/l
RT: 4.558 min Scan# 505
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 43 Resp: 166414
Ion Ratio Lower Upper
43 100
74 23.5 26.4 39.6#





#19
Methyl tert-butyl Ether
Concen: 52.348 ug/l
RT: 5.311 min Scan# 611
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 73 Resp: 423768
Ion Ratio Lower Upper
73 100
57 22.5 16.6 25.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#20
Methylene Chloride
Concen: 50.726 ug/l
RT: 4.793 min Scan# 545
Delta R.T. -0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 84 Resp: 195986
Ion Ratio Lower Upper
84 100
49 132.3 71.4 107.2#
51 38.7 24.5 36.7#
86 65.9 48.7 73.1





#21
trans-1,2-Dichloroethene
Concen: 51.596 ug/l
RT: 5.305 min Scan# 61
Delta R.T. -0.006 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion: 96 Resp: 174930
Ion Ratio Lower Upper
96 100
61 144.4 89.3 133.9
98 60.5 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#22
Diisopropyl ether
Concen: 51.610 ug/l
RT: 6.216 min Scan# 787
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 45 Resp: 599457
Ion Ratio Lower Upper
45 100
43 52.7 40.0 60.0
87 24.8 28.2 42.4
59 11.4 11.0 16.4





#23
 Vinyl Acetate
 Concen: 255.796 ug/l m
 RT: 6.152 min Scan# 776
 Delta R.T. 0.000 min
 Lab File: VD076395.D
 Acq: 12 Jun 2023 15:48




Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD061223




Tgt Ion: 43 Resp: 1276059
 Ion Ratio Lower Upper
 43 100
 86 9.7 12.5 18.7

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 06/13/2023
 Supervised By :Mahesh Dadoda 06/13/2023




#24
 1,1-Dichloroethane
 Concen: 52.327 ug/l
 RT: 6.105 min Scan# 768
 Delta R.T. 0.000 min
 Lab File: VD076395.D
 Acq: 12 Jun 2023 15:48




Tgt Ion: 63 Resp: 337835
 Ion Ratio Lower Upper
 63 100
 98 6.5 4.5 13.7
 100 4.1 2.5 7.5





#25
2-Butanone
Concen: 264.192 ug/l
RT: 7.081 min Scan# 91
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 43 Resp: 30654
Ion Ratio Lower Upper
43 100
72 22.6 30.7 46.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#26
2,2-Dichloropropane
Concen: 52.098 ug/l
RT: 7.069 min Scan# 932
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48



Tgt Ion: 77 Resp: 302847
Ion Ratio Lower Upper
77 100
97 23.0 13.2 39.6





#27

cis-1,2-Dichloroethene

Concen: 51.961 ug/l

RT: 7.075 min Scan# 91

Delta R.T. 0.000 min

Lab File: VD076395.D


Acq: 12 Jun 2023 15:48

Instrument :

MSVOA_D

ClientSampleId :

ICVVD061223



Tgt Ion: 96 Resp: 209710

Ion Ratio Lower Upper

96 100

61 148.9 0.0 239.2




98 64.3 0.0 128.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/13/2023

Supervised By :Mahesh Dadoda 06/13/2023



#28

Bromochloromethane

Concen: 51.276 ug/l

RT: 7.422 min Scan# 992

Delta R.T. 0.000 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48




Tgt Ion: 49 Resp: 91344


Ion Ratio Lower Upper

49 100

129 2.1 0.0 6.0


130 71.7 83.2 124.8#





#29
Tetrahydrofuran
Concen: 257.795 ug/l
RT: 7.446 min Scan# 996
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion: 42 Resp: 190835
Ion Ratio Lower Upper
42 100
72 39.2 46.5 69.7
71 37.2 44.5 66.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#30
Chloroform
Concen: 51.821 ug/l
RT: 7.593 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 83 Resp: 333456
Ion Ratio Lower Upper
83 100
85 64.4 53.4 80.2





#31
Cyclohexane
Concen: 52.302 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion: 56 Resp: 282400
Ion Ratio Lower Upper
56 100
69 31.9 28.6 43.0
84 84.0 84.8 127.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#32
1,1,1-Trichloroethane
Concen: 54.131 ug/l
RT: 7.787 min Scan# 1054
Delta R.T. -0.006 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 97 Resp: 297474
Ion Ratio Lower Upper
97 100
99 63.1 52.0 78.0
61 43.2 33.8 50.8





#33
1,2-Dichloroethane-d4
Concen: 48.861 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 65 Resp: 16489
Ion Ratio Lower Upper
65 100
67 53.3 0.0 115.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.775 min Scan# 1222
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48



Tgt Ion:114 Resp: 499251
Ion Ratio Lower Upper
114 100
63 20.6 0.0 33.4
88 15.4 0.0 26.2





#35

Dibromofluoromethane

Concen: 51.365 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48

Instrument :

MSVOA_D

ClientSampleId :

ICVVD061223

Tgt Ion: 113 Resp: 16574





Ion	Ratio	Lower	Upper
113	100		
111	102.4	84.7	127.1
192	20.1	16.8	25.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/13/2023

Supervised By :Mahesh Dadoda 06/13/2023



#36

1,1-Dichloropropene

Concen: 53.742 ug/l

RT: 8.005 min Scan# 1091




Delta R.T. 0.000 min


Lab File: VD076395.D

Acq: 12 Jun 2023 15:48

Tgt Ion: 75 Resp: 253110

Ion	Ratio	Lower	Upper
75	100		
110	35.3	19.9	59.6
77	31.5	24.7	37.1





#37

Ethyl Acetate

Concen: 52.760 ug/l

RT: 7.169 min Scan# 949

Delta R.T. 0.000 min

Lab File: VD076395.D


Acq: 12 Jun 2023 15:48

Instrument :

MSVOA_D

ClientSampleId :

ICVVD061223






Tgt Ion:	43	Resp:	130699
Ion Ratio	Lower	Upper	
43	100		
61	13.3	14.9	22.3
70	10.5	10.0	15.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/13/2023

Supervised By :Mahesh Dadoda 06/13/2023



#38

Carbon Tetrachloride

Concen: 56.009 ug/l




RT: 7.987 min Scan# 1088


Delta R.T. 0.000 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48

Tgt Ion:	117	Resp:	244912
Ion Ratio	Lower	Upper	
117	100		
119	94.5	78.0	117.0
121	30.4	24.2	36.2





#39

Methylcyclohexane

Concen: 55.949 ug/l

RT: 9.269 min Scan# 1306

Delta R.T. 0.000 min

Lab File: VD076395.D


Acq: 12 Jun 2023 15:48

Instrument :

MSVOA_D

ClientSampleId :

ICVVD061223






Tgt Ion:	83	Resp:	30341
Ion	Ratio	Lower	Upper
83	100		
55	74.7	49.4	74.0
98	45.4	38.4	57.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/13/2023

Supervised By :Mahesh Dadoda 06/13/2023



#40

Benzene

Concen: 52.673 ug/l

RT: 8.252 min Scan# 1133

Delta R.T. 0.000 min

Lab File: VD076395.D




Acq: 12 Jun 2023 15:48


Tgt Ion: 78 Resp: 721192

Ion Ratio Lower Upper

78 100


77 24.6 19.4 29.0





#41
Methacrylonitrile
Concen: 43.461 ug/l
RT: 7.399 min Scan# 91
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 41 Resp: 68874
Ion Ratio Lower Upper
41 100
39 64.2 34.0 51.0
67 76.8 77.0 115.4
52 33.1 27.9 41.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#42
1,2-Dichloroethane
Concen: 50.579 ug/l
RT: 8.322 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 62 Resp: 206225
Ion Ratio Lower Upper
62 100
98 10.0 0.0 25.4





#43
Isopropyl Acetate
Concen: 52.108 ug/l
RT: 8.357 min Scan# 1151
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument : MSVOA_D
ClientSampleId : ICVVD061223



Tgt Ion: 43 Resp: 270638
Ion Ratio Lower Upper
43 100
61 27.9 31.4 47.0
87 11.8 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#44
Trichloroethene
Concen: 54.030 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48

Tgt Ion:130 Resp: 200542
Ion Ratio Lower Upper
130 100
95 97.8 0.0 179.2





#45

1,2-Dichloropropane

Concen: 53.138 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48

Instrument :

MSVOA_D

ClientSampleId :

ICVVD061223

Tgt Ion: 63 Resp: 193374

Ion Ratio Lower Upper

63 100





65 30.2 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/13/2023

Supervised By :Mahesh Dadoda 06/13/2023



#46

Dibromomethane

Concen: 51.658 ug/l

RT: 9.387 min Scan# 1326

Delta R.T. -0.006 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48




Tgt Ion: 93 Resp: 103744


Ion Ratio Lower Upper

93 100

95 83.5 67.7 101.5


174 98.9 85.4 128.2





#47
Bromodichloromethane
Concen: 52.595 ug/l
RT: 9.581 min Scan# 1359
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 83 Resp: 261658
Ion Ratio Lower Upper
83 100
85 65.2 50.2 75.4
127 9.6 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#48
Methyl methacrylate
Concen: 51.378 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 41 Resp: 127898
Ion Ratio Lower Upper
41 100
69 78.1 102.2 153.2#
39 53.6 41.4 62.2





#49
1,4-Dioxane
Concen: 1081.722 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion: 88 Resp: 25779
Ion Ratio Lower Upper
88 100
43 34.0 19.0 28.6
58 69.8 43.5 65.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#50
Toluene-d8
Concen: 50.142 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 98 Resp: 613133
Ion Ratio Lower Upper
98 100
100 65.7 50.3 75.5





#51
4-Methyl-2-Pentanone
Concen: 261.041 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion: 43 Resp: 673394
Ion Ratio Lower Upper
43 100
58 36.8 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#52
Toluene
Concen: 52.624 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 92 Resp: 455491
Ion Ratio Lower Upper
92 100
91 175.6 143.0 214.4





#53
t-1,3-Dichloropropene
Concen: 52.798 ug/l
RT: 10.552 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 75 Resp: 268654
Ion Ratio Lower Upper
75 100
77 33.6 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#54
cis-1,3-Dichloropropene
Concen: 52.312 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 75 Resp: 306106
Ion Ratio Lower Upper
75 100
77 30.6 25.5 38.3
39 51.7 22.1 33.1#





#55
1,1,2-Trichloroethane
Concen: 52.397 ug/l
RT: 10.734 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 97 Resp: 14055
Ion Ratio Lower Upper
97 100
83 86.1 65.8 98.6
85 56.0 45.7 68.5
99 59.3 46.2 69.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#56
Ethyl methacrylate
Concen: 52.461 ug/l
RT: 10.599 min Scan# 1532
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48



Tgt Ion: 69 Resp: 195432
Ion Ratio Lower Upper
69 100
41 70.8 39.1 58.7#
39 40.6 20.0 30.0#





#57

1,3-Dichloropropane

Concen: 52.421 ug/l

RT: 10.881 min Scan# 1580

Delta R.T. 0.000 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48

Instrument :

MSVOA_D

ClientSampleId :

ICVVD061223

Tgt Ion: 76 Resp: 24461

Ion Ratio Lower Upper

76 100





78 32.0 26.4 39.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/13/2023

Supervised By :Mahesh Dadoda 06/13/2023



#58

2-Chloroethyl Vinyl ether

Concen: 242.448 ug/l

RT: 9.869 min Scan# 1408

Delta R.T. 0.000 min

Lab File: VD076395.D




Acq: 12 Jun 2023 15:48


Tgt Ion: 63 Resp: 367895

Ion Ratio Lower Upper

63 100


106 27.5 25.3 37.9





#59
2-Hexanone
Concen: 269.321 ug/l
RT: 10.922 min Scan# 1587
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 43 Resp: 486994
Ion Ratio Lower Upper
43 100
58 52.1 34.1 102.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#60
Dibromochloromethane
Concen: 53.284 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion:129 Resp: 176851
Ion Ratio Lower Upper
129 100
127 77.0 37.1 111.5





#61
1,2-Dibromoethane
Concen: 52.177 ug/l
RT: 11.181 min Scan# 1631
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion:107 Resp: 135789
Ion Ratio Lower Upper
107 100
109 96.1 75.6 113.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#62
4-Bromofluorobenzene
Concen: 52.163 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 95 Resp: 218821
Ion Ratio Lower Upper
95 100
174 82.7 0.0 201.8
176 80.7 0.0 191.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion:117 Resp: 442834
Ion Ratio Lower Upper
117 100
82 53.7 38.3 57.5
119 32.0 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#64
Tetrachloroethene
Concen: 53.752 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion:164 Resp: 163414
Ion Ratio Lower Upper
164 100
166 124.7 99.2 148.8
129 94.7 68.4 102.6
131 93.0 63.8 95.8





#65
Chlorobenzene
Concen: 53.244 ug/l
RT: 11.604 min Scan# 1703
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion:112 Resp: 488732
Ion Ratio Lower Upper
112 100
114 31.8 25.1 37.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#66
1,1,1,2-Tetrachloroethane
Concen: 52.852 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion:131 Resp: 181398
Ion Ratio Lower Upper
131 100
133 94.6 48.5 145.5
119 67.0 35.6 106.9





#67
Ethyl Benzene
Concen: 53.773 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 91 Resp: 90910
Ion Ratio Lower Upper
91 100
106 30.6 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#68
m/p-Xylenes
Concen: 107.162 ug/l
RT: 11.793 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion:106 Resp: 685490
Ion Ratio Lower Upper
106 100
91 201.9 147.0 220.6





#69
o-Xylene
Concen: 53.488 ug/l
RT: 12.122 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion:106 Resp: 328679
Ion Ratio Lower Upper
106 100
91 213.4 97.9 293.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#70
Styrene
Concen: 53.212 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion:104 Resp: 557874
Ion Ratio Lower Upper
104 100
78 50.2 33.2 49.8#
103 55.5 43.6 65.4





#71
Bromoform
Concen: 53.523 ug/l
RT: 12.298 min Scan# 1821
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion:173 Resp: 10714
Ion Ratio Lower Upper
173 100
175 47.9 24.6 73.8
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion:152 Resp: 208313
Ion Ratio Lower Upper
152 100
115 59.3 27.1 81.3
150 177.6 0.0 353.6





#73
Isopropylbenzene
Concen: 54.675 ug/l
RT: 12.422 min Scan# 1842
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument : MSVOA_D
ClientSampleId : ICVVD061223



Tgt Ion:105 Resp: 86135
Ion Ratio Lower Upper
105 100
120 26.7 14.3 42.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#74
N-ethyl acetate
Concen: 53.477 ug/l
RT: 12.234 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 43 Resp: 256923
Ion Ratio Lower Upper
43 100
70 39.4 45.0 67.4#
55 21.1 23.8 35.6#
61 22.9 22.1 33.1





#75
1,1,2,2-Tetrachloroethane
Concen: 52.226 ug/l
RT: 12.675 min Scan# 1885
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion: 83 Resp: 16247
Ion Ratio Lower Upper
83 100
131 10.1 5.1 15.3
85 64.3 32.6 97.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#76
1,2,3-Trichloropropane
Concen: 47.649 ug/l m
RT: 12.728 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 75 Resp: 92796
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 53.581 ug/l
RT: 12.704 min Scan# 1889
Delta R.T. 0.006 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion:156 Resp: 199160
Ion Ratio Lower Upper
156 100
77 186.4 76.8 230.4
158 96.6 48.6 145.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#78
n-propylbenzene
Concen: 54.981 ug/l
RT: 12.763 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 91 Resp: 1055790
Ion Ratio Lower Upper
91 100
120 22.8 12.4 37.2





#79
2-Chlorotoluene
Concen: 53.596 ug/l
RT: 12.851 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion: 91 Resp: 582397
Ion Ratio Lower Upper
91 100
126 34.8 19.4 58.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#80
1,3,5-Trimethylbenzene
Concen: 54.147 ug/l
RT: 12.904 min Scan# 1924
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion:105 Resp: 712001
Ion Ratio Lower Upper
105 100
120 48.6 26.4 79.2





#81
trans-1,4-Dichloro-2-butene
Concen: 49.999 ug/l
RT: 12.469 min Scan# 1850
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion: 75 Resp: 54384
Ion Ratio Lower Upper
75 100
53 91.0 63.0 94.4
89 47.1 37.9 56.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#82
4-Chlorotoluene
Concen: 54.054 ug/l
RT: 12.945 min Scan# 1931
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion: 91 Resp: 615102
Ion Ratio Lower Upper
91 100
126 34.2 19.4 58.4





#83
tert-Butylbenzene
Concen: 54.470 ug/l
RT: 13.169 min Scan# 1969
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion:119 Resp: 614650
Ion Ratio Lower Upper
119 100
91 65.6 29.3 87.8
134 27.2 14.1 42.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#84
1,2,4-Trimethylbenzene
Concen: 54.381 ug/l
RT: 13.210 min Scan# 1976
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48



Tgt Ion:105 Resp: 706517
Ion Ratio Lower Upper
105 100
120 45.5 24.3 72.9





#85

sec-Butylbenzene

Concen: 55.290 ug/l

RT: 13.345 min Scan# 1999

Delta R.T. 0.000 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48

Instrument :

MSVOA_D

ClientSampleId :

ICVVD061223

Tgt Ion:105 Resp: 91828

Ion Ratio Lower Upper

105 100





134 20.4 10.9 32.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/13/2023

Supervised By :Mahesh Dadoda 06/13/2023



#86

p-Isopropyltoluene

Concen: 55.645 ug/l

RT: 13.463 min Scan# 2019

Delta R.T. 0.000 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48




Tgt Ion:119 Resp: 778200


Ion Ratio Lower Upper

119 100

134 26.4 14.2 42.6


91 23.9 10.5 31.5





#87
1,3-Dichlorobenzene
Concen: 54.688 ug/l
RT: 13.457 min Scan# 2018
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion:146 Resp: 39340
Ion Ratio Lower Upper
146 100
111 39.6 17.9 53.8
148 63.1 32.3 96.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#88
1,4-Dichlorobenzene
Concen: 54.429 ug/l
RT: 13.540 min Scan# 2032
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48



Tgt Ion:146 Resp: 386938
Ion Ratio Lower Upper
146 100
111 38.7 17.4 52.3
148 63.4 31.7 95.1





#89

n-Butylbenzene

Concen: 56.067 ug/l

RT: 13.787 min Scan# 2074

Delta R.T. 0.000 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48

Instrument :

MSVOA_D

ClientSampleId :

ICVVD061223





Tgt Ion:	91	Resp:	753349
Ion Ratio	Lower	Upper	
91	100		
92	53.5	26.5	79.5
134	25.8	13.8	41.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/13/2023

Supervised By :Mahesh Dadoda 06/13/2023



#90

Hexachloroethane

Concen: 55.048 ug/l




RT: 14.051 min Scan# 2119


Delta R.T. -0.006 min

Lab File: VD076395.D

Acq: 12 Jun 2023 15:48


Tgt Ion:	117	Resp:	142068
Ion Ratio	Lower	Upper	
117	100		
201	75.7	38.5	115.5





#91
1,2-Dichlorobenzene
Concen: 53.782 ug/l
RT: 13.834 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223



Tgt Ion:146 Resp: 34025
Ion Ratio Lower Upper
146 100
111 40.5 18.3 54.8
148 63.9 31.6 94.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.841 ug/l
RT: 14.445 min Scan# 2186
Delta R.T. -0.006 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48


Tgt Ion: 75 Resp: 25471
Ion Ratio Lower Upper
75 100
155 92.1 47.9 143.8
157 114.9 59.9 179.7





#93
1,2,4-Trichlorobenzene
Concen: 54.974 ug/l
RT: 15.098 min Scan# 2297
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion:180 Resp: 232774
Ion Ratio Lower Upper
180 100
182 95.2 48.5 145.6
145 30.3 14.2 42.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023




#94
Hexachlorobutadiene
Concen: 56.157 ug/l
RT: 15.204 min Scan# 2315
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Tgt Ion:225 Resp: 119628
Ion Ratio Lower Upper
225 100
223 62.8 31.6 94.8
227 63.9 31.7 95.1





#95
Naphthalene
Concen: 53.212 ug/l
RT: 15.334 min Scan# 2337
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48




Instrument :
MSVOA_D
ClientSampleId :
ICVVD061223




Tgt Ion:128 Resp: 452440
Ion Ratio Lower Upper
128 100
127 13.3 10.5 15.7
129 11.3 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/13/2023
Supervised By :Mahesh Dadoda 06/13/2023



#96
1,2,3-Trichlorobenzene
Concen: 54.969 ug/l
RT: 15.522 min Scan# 2369
Delta R.T. 0.000 min
Lab File: VD076395.D
Acq: 12 Jun 2023 15:48



Tgt Ion:180 Resp: 201023
Ion Ratio Lower Upper
180 100
182 94.8 47.3 141.9
145 32.2 15.1 45.3

