

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061322\
 Data File : VD073542.D
 Acq On : 13 Jun 2022 13:15
 Operator : VA/SY
 Sample : N3233-02
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CR-2

Quant Time: Jun 13 15:32:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	221591	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	399484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	454653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	210789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	109391	46.002	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.000%
35) Dibromofluoromethane	7.908	113	134084	52.196	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.400%
50) Toluene-d8	10.332	98	500681	51.550	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.100%
62) 4-Bromofluorobenzene	12.620	95	206517	58.572	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	117.140%
Target Compounds						
16) Acetone	4.162	43	9642	23.028	ug/l	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061322\
Data File : VD073542.D
Acq On : 13 Jun 2022 13:15
Operator : VA/SY
Sample : N3233-02
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CR-2

Quant Time: Jun 13 15:32:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

