

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061322\
 Data File : VD073543.D
 Acq On : 13 Jun 2022 13:45
 Operator : VA/SY
 Sample : N3282-01
 Misc : 4.72G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 125

Quant Time: Jun 13 15:33:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	210804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	404306	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	474012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	251591	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	127761	56.476	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.960%
35) Dibromofluoromethane	7.908	113	138398	53.233	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.460%
50) Toluene-d8	10.332	98	509807	51.864	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.720%
62) 4-Bromofluorobenzene	12.620	95	245637	68.836	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	137.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061322\
Data File : VD073543.D
Acq On : 13 Jun 2022 13:45
Operator : VA/SY
Sample : N3282-01
Misc : 4.72G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
125

Quant Time: Jun 13 15:33:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

