

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061322\
 Data File : VD073546.D
 Acq On : 13 Jun 2022 15:13
 Operator : VA/SY
 Sample : N3234-01RE
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RB-21499RE

Quant Time: Jun 14 01:28:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	30428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	50667	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	68136	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	35526	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	24995	76.547	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.100%
35) Dibromofluoromethane	7.908	113	22518	69.114	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	138.220%
50) Toluene-d8	10.337	98	63748	51.750	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.500%
62) 4-Bromofluorobenzene	12.620	95	30271	67.691	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	135.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061322\
Data File : VD073546.D
Acq On : 13 Jun 2022 15:13
Operator : VA/SY
Sample : N3234-01RE
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RB-21499RE

Quant Time: Jun 14 01:28:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

