


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061322\
Data File : VD073536.D
Acq On : 13 Jun 2022 09:15
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Title : SW846 8260
Last Update : Thu Jun 09 12:33:32 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	27056	PASS
75	95	30	60	50.8	78336	PASS
95	95	100	100	100.0	154133	PASS
96	95	5	9	6.4	9896	PASS
173	174	0.00	2	0.9	1103	PASS
174	95	50	100	78.2	120504	PASS
175	174	5	9	7.6	9183	PASS
176	174	95	101	96.7	116515	PASS
177	176	5	9	7.0	8202	PASS