

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061323\
 Data File : VD076403.D
 Acq On : 13 Jun 2023 14:18
 Operator : KP/SY
 Sample : 03074-01RE
 Misc : 6.92G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GT-01RE

Quant Time: Jun 14 05:46:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061223S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 13 02:01:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	127571	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	239461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	255091	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.521	152	128235	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	105176	71.379	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 142.760%	
35) Dibromofluoromethane	7.804	113	123103	79.541	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 159.080%#	
50) Toluene-d8	10.269	98	321157	54.758	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 109.520%	
62) 4-Bromofluorobenzene	12.574	95	147864	73.489	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 146.980%#	
Target Compounds						
						Qvalue
16) Acetone	4.028	43	7569	23.367	ug/l	95
20) Methylene Chloride	4.793	84	34383	15.968	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061323\
Data File : VD076403.D
Acq On : 13 Jun 2023 14:18
Operator : KP/SY
Sample : 03074-01RE
Misc : 6.92G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GT-01RE

Quant Time: Jun 14 05:46:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061223S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 13 02:01:34 2023
Response via : Initial Calibration

