

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD061419\
 Data File : VD062704.D
 Acq On : 14 Jun 2019 11:58
 Operator : VA/AP
 Sample : VD0614SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0614SBL01

Quant Time: Jun 15 05:02:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 00:02:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	766933	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.23	114	1060650	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	819221	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	480207	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	564097	50.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.94%	
35) Dibromofluoromethane	6.94	113	576976	51.58	ug/l	0.02
Spiked Amount	50.000		Recovery	=	103.16%	
50) Toluene-d8	10.41	98	1203062	51.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.26%	
62) 4-Bromofluorobenzene	13.56	95	562789	51.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.52%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD061419\
Data File : VD062704.D
Acq On : 14 Jun 2019 11:58
Operator : VA/AP
Sample : VD0614SBL01
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0614SBL01

Quant Time: Jun 15 05:02:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060719S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 00:02:21 2019
Response via : Initial Calibration

