

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061421\
 Data File : VD069424.D
 Acq On : 14 Jun 2021 13:58
 Operator : VA/SY
 Sample : M2678-07
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 0241-AMND-COMP-F(CL-05)

Quant Time: Jun 15 07:26:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 12:50:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	247465	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	453756	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	466885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	205197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	161473	57.838	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.680%
35) Dibromofluoromethane	7.920	113	94284	27.742	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	55.480%
50) Toluene-d8	10.338	98	572798	46.080	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.160%
62) 4-Bromofluorobenzene	12.626	95	184289	46.205	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.420%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	33174	71.789	ug/l	100
17) Carbon Disulfide	4.409	76	61601	7.256	ug/l	99
25) 2-Butanone	7.220	43	6471	11.230	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061421\
Data File : VD069424.D
Acq On : 14 Jun 2021 13:58
Operator : VA/SY
Sample : M2678-07
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
0241-AMND-COMP-F(CL-05)

Quant Time: Jun 15 07:26:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 11 12:50:13 2021
Response via : Initial Calibration

