



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061421\
 Data File : VD069417.D
 Acq On : 14 Jun 2021 09:09
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Title : SW846 8260
 Last Update : Fri Jun 11 12:50:13 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	34315	PASS
75	95	30	60	50.5	94307	PASS
95	95	100	100	100.0	186923	PASS
96	95	5	9	6.5	12156	PASS
173	174	0.00	2	1.3	2093	PASS
174	95	50	100	86.4	161507	PASS
175	174	5	9	7.7	12429	PASS
176	174	95	101	95.3	153843	PASS
177	176	5	9	6.9	10618	PASS