

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061422\
 Data File : VD073563.D
 Acq On : 14 Jun 2022 12:09
 Operator : VA/SY
 Sample : N3296-01
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-03-060922

Quant Time: Jun 14 15:43:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	204734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	367927	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	457173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	228571	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	105408	47.977	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.960%
35) Dibromofluoromethane	7.908	113	125342	52.978	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.960%
50) Toluene-d8	10.332	98	470811	52.633	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.260%
62) 4-Bromofluorobenzene	12.620	95	211919	65.259	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	130.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061422\
Data File : VD073563.D
Acq On : 14 Jun 2022 12:09
Operator : VA/SY
Sample : N3296-01
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-03-060922

Quant Time: Jun 14 15:43:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

