

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061422\
 Data File : VD073564.D
 Acq On : 14 Jun 2022 12:38
 Operator : VA/SY
 Sample : N3328-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BU-02-061022

Quant Time: Jun 15 01:35:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	196605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	355077	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	432657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	230969	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	102966	48.803	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.600%
35) Dibromofluoromethane	7.908	113	119310	52.254	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.500%
50) Toluene-d8	10.332	98	451428	52.292	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.580%
62) 4-Bromofluorobenzene	12.620	95	208012	66.374	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	132.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061422\
Data File : VD073564.D
Acq On : 14 Jun 2022 12:38
Operator : VA/SY
Sample : N3328-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BU-02-061022

Quant Time: Jun 15 01:35:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

