

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061422\
 Data File : VD073567.D
 Acq On : 14 Jun 2022 14:05
 Operator : VA/SY
 Sample : N3287-01
 Misc : 4.72G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1

Quant Time: Jun 15 01:36:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	185800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	344755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	441415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	241176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	104885	52.603	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.200%	
35) Dibromofluoromethane	7.902	113	117611	53.052	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.100%	
50) Toluene-d8	10.332	98	446677	53.291	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.580%	
62) 4-Bromofluorobenzene	12.620	95	228610	75.130	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 150.260%#	
Target Compounds						Qvalue
20) Methylene Chloride	4.961	84	26822	9.550	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061422\
Data File : VD073567.D
Acq On : 14 Jun 2022 14:05
Operator : VA/SY
Sample : N3287-01
Misc : 4.72G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1

Quant Time: Jun 15 01:36:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

