

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061422\
 Data File : VD073578.D
 Acq On : 14 Jun 2022 19:33
 Operator : VA/SY
 Sample : N3331-01
 Misc : 7.69G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-222(26-28)

Quant Time: Jun 15 01:38:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	267653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	484673	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	635544	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	369505	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	134896	46.965	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.920%
35) Dibromofluoromethane	7.908	113	156989	50.371	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.740%
50) Toluene-d8	10.332	98	628286	53.319	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.640%
62) 4-Bromofluorobenzene	12.620	95	348318	81.425	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	162.860%#
Target Compounds						
					Qvalue	
16) Acetone	4.155	43	8809	17.418	ug/l	99
20) Methylene Chloride	4.950	84	33468	7.804	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061422\
Data File : VD073578.D
Acq On : 14 Jun 2022 19:33
Operator : VA/SY
Sample : N3331-01
Misc : 7.69G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-222(26-28)

Quant Time: Jun 15 01:38:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

