

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061521\
 Data File : VD069448.D
 Acq On : 15 Jun 2021 19:04
 Operator : VA/SY
 Sample : M2709-11
 Misc : 5.090G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 0250-AMND-COMP-O(CL-23)

Quant Time: Jun 16 02:02:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 12:50:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	312983	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	627945	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	698105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	347543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	188849	53.484	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery	= 106.960%		
35) Dibromofluoromethane	7.914	113	133934	28.476	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery	= 56.960%		
50) Toluene-d8	10.338	98	873056	50.752	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery	= 101.500%		
62) 4-Bromofluorobenzene	12.626	95	362254	65.631	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery	= 131.260%		
Target Compounds						Qvalue
16) Acetone	4.156	43	45362	77.615	ug/l	94
17) Carbon Disulfide	4.403	76	153313	14.279	ug/l	97
39) Methylcyclohexane	9.349	83	26947	4.097	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061521\
Data File : VD069448.D
Acq On : 15 Jun 2021 19:04
Operator : VA/SY
Sample : M2709-11
Misc : 5.090G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
0250-AMND-COMP-O(CL-23)

Quant Time: Jun 16 02:02:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 11 12:50:13 2021
Response via : Initial Calibration

