

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061521\
 Data File : VD069452.D
 Acq On : 15 Jun 2021 20:57
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 16 02:03:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 12:50:13 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.498	0.478	4.0	108	0.00
3 P	Chloromethane	0.498	0.449	9.8	100	0.00
4 C	Vinyl Chloride	0.603	0.540	10.4#	99	0.00
5 T	Bromomethane	0.491	0.440	10.4	106	0.00
6 T	Chloroethane	0.415	0.392	5.5	105	0.00
7 T	Trichlorofluoromethane	0.986	0.938	4.9	109	0.00
8 T	Diethyl Ether	0.247	0.278	-12.6	118	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.586	0.565	3.6	110	0.00
10 T	Methyl Iodide	0.536	0.633	-18.1	117	0.00
11 T	Tert butyl alcohol	0.029	0.029	0.0	114	0.00
12 CM	1,1-Dichloroethene	0.530	0.533	-0.6#	109	0.00
13 T	Acrolein	0.037	0.035	5.4	99	0.00
14 T	Allyl chloride	0.624	0.694	-11.2	121	0.00
15 T	Acrylonitrile	0.095	0.106	-11.6	121	0.00
16 T	Acetone	0.093	0.081	12.9	91	0.00
17 T	Carbon Disulfide	1.715	1.741	-1.5	109	0.00
18 T	Methyl Acetate	0.254	0.257	-1.2	128	0.00
19 T	Methyl tert-butyl Ether	0.967	1.151	-19.0	123	0.00
20 T	Methylene Chloride	0.875	0.870	0.6	137	0.00
21 T	trans-1,2-Dichloroethene	0.621	0.665	-7.1	117	0.00
22 T	Diisopropyl ether	1.240	1.465	-18.1	121	0.00
23 T	Vinyl Acetate	0.690	0.844	-22.3	118	0.00
24 P	1,1-Dichloroethane	1.032	1.088	-5.4	118	0.00
25 T	2-Butanone	0.116	0.126	-8.6	112	0.00
26 T	2,2-Dichloropropane	0.910	0.907	0.3	111	0.00
27 T	cis-1,2-Dichloroethene	0.672	0.748	-11.3	120	0.00
28 T	Bromochloromethane	0.378	0.405	-7.1	112	0.00
29 T	Tetrahydrofuran	0.066	0.077	-16.7	118	0.00
30 C	Chloroform	1.150	1.206	-4.9#	118	0.00
31 T	Cyclohexane	0.767	0.755	1.6	108	0.00
32 T	1,1,1-Trichloroethane	1.010	1.025	-1.5	113	0.00
33 S	1,2-Dichloroethane-d4	0.564	0.564	0.0	110	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	119	0.00
35 S	Dibromofluoromethane	0.375	0.360	4.0	112	0.00
36 T	1,1-Dichloropropene	0.521	0.514	1.3	112	0.00
37 T	Ethyl Acetate	0.174	0.183	-5.2	118	0.00
38 T	Carbon Tetrachloride	0.564	0.544	3.5	113	0.00
39 T	Methylcyclohexane	0.524	0.531	-1.3	111	0.00
40 TM	Benzene	1.515	1.540	-1.7	117	0.00
41 T	Methacrylonitrile	0.087	0.104	-19.5	141	0.00
42 TM	1,2-Dichloroethane	0.416	0.417	-0.2	118	0.00
43 T	Isopropyl Acetate	0.329	0.346	-5.2	118	0.00
44 TM	Trichloroethene	0.423	0.424	-0.2	117	0.00
45 C	1,2-Dichloropropane	0.370	0.377	-1.9#	118	0.00
46 T	Dibromomethane	0.212	0.219	-3.3	121	0.00
47 T	Bromodichloromethane	0.546	0.551	-0.9	118	0.00
48 T	Methyl methacrylate	0.150	0.162	-8.0	118	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061521\
 Data File : VD069452.D
 Acq On : 15 Jun 2021 20:57
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 16 02:03:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 12:50:13 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	109	0.00
50 S	Toluene-d8	1.370	1.354	1.2	111	0.00
51 T	4-Methyl-2-Pentanone	0.168	0.183	-8.9	116	0.00
52 CM	Toluene	0.962	1.022	-6.2#	119	0.00
53 T	t-1,3-Dichloropropene	0.468	0.500	-6.8	121	0.00
54 T	cis-1,3-Dichloropropene	0.563	0.582	-3.4	118	0.00
55 T	1,1,2-Trichloroethane	0.293	0.302	-3.1	120	0.00
56 T	Ethyl methacrylate	0.296	0.350	-18.2	125	0.00
57 T	1,3-Dichloropropane	0.477	0.498	-4.4	120	0.00
58 T	2-Chloroethyl Vinyl ether	0.149	0.162	-8.7	112	0.00
59 T	2-Hexanone	0.112	0.120	-7.1	113	0.00
60 T	Dibromochloromethane	0.373	0.384	-2.9	119	0.00
61 T	1,2-Dibromoethane	0.281	0.294	-4.6	121	0.00
62 S	4-Bromofluorobenzene	0.439	0.446	-1.6	115	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	121	0.00
64 T	Tetrachloroethene	0.387	0.371	4.1	116	0.00
65 PM	Chlorobenzene	1.100	1.119	-1.7	120	0.00
66 T	1,1,1,2-Tetrachloroethane	0.422	0.419	0.7	118	0.00
67 C	Ethyl Benzene	1.844	1.943	-5.4#	118	0.00
68 T	m/p-Xylenes	0.733	0.768	-4.8	115	0.00
69 T	o-Xylene	0.648	0.707	-9.1	119	0.00
70 T	Styrene	1.141	1.253	-9.8	119	0.00
71 P	Bromoform	0.219	0.219	0.0	117	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
73 T	Isopropylbenzene	3.515	3.898	-10.9	117	0.00
74 T	N-amyl acetate	0.631	0.698	-10.6	116	0.00
75 P	1,1,2,2-Tetrachloroethane	0.675	0.703	-4.1	118	0.00
76 T	1,2,3-Trichloropropane	0.507	0.497	2.0	108	0.00
77 T	Bromobenzene	0.881	0.938	-6.5	118	0.00
78 T	n-propylbenzene	4.314	4.692	-8.8	113	0.00
79 T	2-Chlorotoluene	2.452	2.682	-9.4	116	0.00
80 T	1,3,5-Trimethylbenzene	2.963	3.329	-12.4	116	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.202	-6.3	113	0.00
82 T	4-Chlorotoluene	2.610	2.842	-8.9	117	0.00
83 T	tert-Butylbenzene	2.523	2.875	-14.0	121	0.00
84 T	1,2,4-Trimethylbenzene	2.963	3.352	-13.1	117	0.00
85 T	sec-Butylbenzene	3.709	3.948	-6.4	112	0.00
86 T	p-Isopropyltoluene	3.222	3.523	-9.3	111	0.00
87 T	1,3-Dichlorobenzene	1.765	1.840	-4.2	115	0.00
88 T	1,4-Dichlorobenzene	1.758	1.811	-3.0	117	0.00
89 T	n-Butylbenzene	2.997	3.221	-7.5	112	0.00
90 T	Hexachloroethane	0.627	0.604	3.7	108	0.00
91 T	1,2-Dichlorobenzene	1.540	1.587	-3.1	114	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.102	0.104	-2.0	120	0.00
93 T	1,2,4-Trichlorobenzene	0.896	0.986	-10.0	117	0.00
94 T	Hexachlorobutadiene	0.560	0.455	18.8	87	0.00
95 T	Naphthalene	1.515	1.781	-17.6	120	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061521\
Data File : VD069452.D
Acq On : 15 Jun 2021 20:57
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jun 16 02:03:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 11 12:50:13 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.778	0.865	-11.2	116	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6