

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061522\
 Data File : VD073586.D
 Acq On : 15 Jun 2022 12:01
 Operator : VA/SY
 Sample : N3287-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1

Quant Time: Jun 16 06:02:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	300092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	506507	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	482883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	231779	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	175518	54.502	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.000%
35) Dibromofluoromethane	7.903	113	170780	52.434	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.860%
50) Toluene-d8	10.332	98	611798	49.681	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.360%
62) 4-Bromofluorobenzene	12.620	95	218149	48.798	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.600%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061522\
Data File : VD073586.D
Acq On : 15 Jun 2022 12:01
Operator : VA/SY
Sample : N3287-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1

Quant Time: Jun 16 06:02:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1029
Delta R.T. 0.000 min
Lab File: VD073586.D
Acq: 15 Jun 2022 12:01




Instrument :
MSVOA_D
ClientSampleId :
SB-1


Tgt Ion:168 Resp: 300092
Ion Ratio Lower Upper
168 100
99 59.3 49.8 74.6



#33
1,2-Dichloroethane-d4
Concen: 54.502 ug/l
RT: 8.320 min Scan# 1088
Delta R.T. 0.000 min
Lab File: VD073586.D
Acq: 15 Jun 2022 12:01

Tgt Ion: 65 Resp: 175518
Ion Ratio Lower Upper
65 100
67 51.0 0.0 101.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.856 min Scan# 1179

Delta R.T. 0.001 min

Lab File: VD073586.D

Acq: 15 Jun 2022 12:01

Instrument :

MSVOA_D

ClientSampleId :

SB-1





Tgt Ion:114 Resp: 506507

Ion Ratio Lower Upper

114 100

63 19.2 0.0 46.6

88 16.5 0.0 33.2



#35

Dibromofluoromethane

Concen: 52.434 ug/l

RT: 7.903 min Scan# 1017

Delta R.T. -0.005 min

Lab File: VD073586.D

Acq: 15 Jun 2022 12:01




Tgt Ion:113 Resp: 170780


Ion Ratio Lower Upper

113 100

111 99.9 83.8 125.6

192 19.0 14.6 21.8









#50
Toluene-d8
Concen: 49.681 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD073586.D
Acq: 15 Jun 2022 12:01




Instrument :
MSVOA_D
ClientSampleId :
SB-1


Tgt Ion: 98 Resp: 611798
Ion Ratio Lower Upper
98 100
100 66.0 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 48.798 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. 0.000 min
Lab File: VD073586.D
Acq: 15 Jun 2022 12:01

Tgt Ion: 95 Resp: 218149
Ion Ratio Lower Upper
95 100
174 80.5 0.0 152.8
176 75.6 0.0 145.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 117
Delta R.T. 0.006 min
Lab File: VD073586.D
Acq: 15 Jun 2022 12:01

Instrument :
MSVOA_D
ClientSampleId :
SB-1

Tgt Ion:117 Resp: 482883
Ion Ratio Lower Upper
117 100
82 53.7 45.3 67.9
119 32.5 26.6 40.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VD073586.D
Acq: 15 Jun 2022 12:01

Tgt Ion:152 Resp: 231779
Ion Ratio Lower Upper
152 100
115 60.4 31.3 93.8
150 157.4 0.0 348.6

