

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061522\
 Data File : VD073601.D
 Acq On : 15 Jun 2022 19:59
 Operator : VA/SY
 Sample : N3337-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-02-061322

Quant Time: Jun 16 06:04:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	368994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	630682	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	583254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	257561	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	202946	51.252	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.500%
35) Dibromofluoromethane	7.908	113	202849	50.018	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.040%
50) Toluene-d8	10.332	98	748993	48.847	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.700%
62) 4-Bromofluorobenzene	12.620	95	248330	44.612	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061522\
Data File : VD073601.D
Acq On : 15 Jun 2022 19:59
Operator : VA/SY
Sample : N3337-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-02-061322

Quant Time: Jun 16 06:04:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1029
Delta R.T. 0.000 min
Lab File: VD073601.D
Acq: 15 Jun 2022 19:59




Instrument :
MSVOA_D
ClientSampleId :
SU-02-061322


Tgt Ion:168 Resp: 368994
Ion Ratio Lower Upper
168 100
99 57.3 49.8 74.6



#33
1,2-Dichloroethane-d4
Concen: 51.252 ug/l
RT: 8.320 min Scan# 1088
Delta R.T. 0.000 min
Lab File: VD073601.D
Acq: 15 Jun 2022 19:59

Tgt Ion: 65 Resp: 202946
Ion Ratio Lower Upper
65 100
67 52.2 0.0 101.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.855 min Scan# 1179

Delta R.T. 0.000 min

Lab File: VD073601.D

Acq: 15 Jun 2022 19:59

Instrument :

MSVOA_D

ClientSampleId :

SU-02-061322





Tgt Ion:114 Resp: 630682

Ion Ratio Lower Upper

114 100

63 19.8 0.0 46.6

88 16.4 0.0 33.2



#35

Dibromofluoromethane

Concen: 50.018 ug/l

RT: 7.908 min Scan# 1018

Delta R.T. 0.000 min

Lab File: VD073601.D

Acq: 15 Jun 2022 19:59




Tgt Ion:113 Resp: 202849


Ion Ratio Lower Upper

113 100

111 101.6 83.8 125.6

192 19.0 14.6 21.8









#50
Toluene-d8
Concen: 48.847 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD073601.D
Acq: 15 Jun 2022 19:59




Instrument :
MSVOA_D
ClientSampleId :
SU-02-061322


Tgt Ion: 98 Resp: 748993
Ion Ratio Lower Upper
98 100
100 66.2 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 44.612 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VD073601.D
Acq: 15 Jun 2022 19:59

Tgt Ion: 95 Resp: 248330
Ion Ratio Lower Upper
95 100
174 77.5 0.0 152.8
176 74.1 0.0 145.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.632 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD073601.D
Acq: 15 Jun 2022 19:59

Instrument :
MSVOA_D
ClientSampleId :
SU-02-061322

Tgt Ion:117 Resp: 583254
Ion Ratio Lower Upper
117 100
82 55.1 45.3 67.9
119 31.0 26.6 40.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. -0.000 min
Lab File: VD073601.D
Acq: 15 Jun 2022 19:59

Tgt Ion:152 Resp: 257561
Ion Ratio Lower Upper
152 100
115 59.3 31.3 93.8
150 156.1 0.0 348.6

