

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061523\
 Data File : VD076436.D
 Acq On : 15 Jun 2023 13:20
 Operator : KP/SY
 Sample : VIBLK
 Misc : 5.00G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Jun 16 01:42:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM061523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 16 01:37:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	279204	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	241807	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	116986	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	175952	25.878	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.520%
7) Chloroethane-d5	2.799	69	136736	24.870	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.480%
11) 1,1-Dichloroethene-d2	3.917	65	45559	25.669	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	102.680%
21) 2-Butanone-d5	6.981	46	51552	50.062	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.120%
24) Chloroform-d	7.563	84	176604	24.008	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.040%
26) 1,2-Dichloroethane-d4	8.228	65	89892	24.771	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.080%
32) Benzene-d6	8.199	84	384664	26.005	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	104.040%
36) 1,2-Dichloropropane-d6	9.210	67	121408	25.714	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.840%
41) Toluene-d8	10.269	98	355942	25.988	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	103.960%
43) trans-1,3-Dichloroprop...	10.522	79	52518	26.743	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	106.960%
47) 2-Hexanone-d5	10.875	63	38671	52.763	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.520%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	84336	25.107	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.440%
66) 1,2-Dichlorobenzene-d4	13.816	152	111025	25.892	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.560%
Target Compounds						Qvalue
10) 1,1,2-Trichloro-1,2,2-...	3.917	101	1102	0.301	ug/L #	20
16) Methylene chloride	4.793	84	9654	1.989	ug/L	88
39) cis-1,3-Dichloropropene	9.975	75	4369	0.656	ug/L	88
44) trans-1,3-Dichloropropene	10.528	75	2549	0.471	ug/L #	34
64) 1,3-Dichlorobenzene	13.457	146	2712	0.341	ug/L	87
65) 1,4-Dichlorobenzene	13.540	146	3395	0.423	ug/L	90
67) 1,2-Dichlorobenzene	13.834	146	2590	0.365	ug/L #	71
69) 1,3,5-Trichlorobenzene	14.592	180	2610	0.492	ug/L	96
70) 1,2,4-trichlorobenzene	15.104	180	3894	0.824	ug/L	99
71) Naphthalene	15.339	128	10768	1.190	ug/L #	95
72) 1,2,3-Trichlorobenzene	15.522	180	4301	1.048	ug/L #	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061523\
Data File : VD076436.D
Acq On : 15 Jun 2023 13:20
Operator : KP/SY
Sample : VIBLK
Misc : 5.00G/10.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :

Quant Time: Jun 16 01:42:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM061523SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 16 01:37:00 2023
Response via : Initial Calibration









#10
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 0.301 ug/L
RT: 3.917 min Scan# 396
Delta R.T. -0.041 min
Lab File: VD076436.D
Acq: 15 Jun 2023 13:20




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:101 Resp: 1102
Ion Ratio Lower Upper
101 100
85 0.0 35.5 53.3#
151 0.0 58.6 87.8#



#16
Methylene chloride
Concen: 1.989 ug/L
RT: 4.793 min Scan# 545
Delta R.T. -0.006 min
Lab File: VD076436.D
Acq: 15 Jun 2023 13:20

Tgt Ion: 84 Resp: 9654
Ion Ratio Lower Upper
84 100
86 70.0 45.1 83.9
49 121.2 97.2 180.6





#39

cis-1,3-Dichloropropene

Concen: 0.656 ug/L

RT: 9.975 min Scan# 1433

Delta R.T. -0.041 min

Lab File: VD076436.D

Acq: 15 Jun 2023 13:20

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion: 75 Resp: 4369

Ion Ratio Lower Upper

75 100

77 39.0 22.7 42.3



#44

trans-1,3-Dichloropropene

Concen: 0.471 ug/L

RT: 10.528 min Scan# 1520

Delta R.T. -0.024 min

Lab File: VD076436.D




Acq: 15 Jun 2023 13:20


Tgt Ion: 75 Resp: 2549

Ion Ratio Lower Upper

75 100

77 69.2 22.5 41.7#









#64
1,3-Dichlorobenzene
Concen: 0.341 ug/L
RT: 13.457 min Scan# 2019
Delta R.T. -0.006 min
Lab File: VD076436.D
Acq: 15 Jun 2023 13:20




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:146 Resp: 2712
Ion Ratio Lower Upper
146 100
111 45.5 25.6 47.4
148 72.4 44.3 82.3



#65
1,4-Dichlorobenzene
Concen: 0.423 ug/L
RT: 13.540 min Scan# 2032
Delta R.T. -0.000 min
Lab File: VD076436.D
Acq: 15 Jun 2023 13:20

Tgt Ion:146 Resp: 3395
Ion Ratio Lower Upper
146 100
111 27.7 25.7 47.7
148 58.8 45.4 84.2





#67

1,2-Dichlorobenzene

Concen: 0.365 ug/L

RT: 13.834 min Scan# 2082

Delta R.T. -0.000 min


Lab File: VD076436.D

Acq: 15 Jun 2023 13:20




Instrument :

MSVOA_D

ClientSampleId :



Tgt Ion:	146	Resp:	2590
Ion Ratio	Lower	Upper	
146	100		
111	63.3	31.3	46.9#
148	46.6	51.7	77.5#



#69

1,3,5-Trichlorobenzene

Concen: 0.492 ug/L




RT: 14.592 min Scan# 2211


Delta R.T. -0.006 min

Lab File: VD076436.D

Acq: 15 Jun 2023 13:20


Tgt Ion:	180	Resp:	2610
Ion Ratio	Lower	Upper	
180	100		
182	94.4	77.1	115.7
184	26.6	25.0	37.4
145	25.0	23.4	35.0








#70
1,2,4-trichlorobenzene
Concen: 0.824 ug/L
RT: 15.104 min Scan# 21
Delta R.T. 0.006 min
Lab File: VD076436.D
Acq: 15 Jun 2023 13:20


Instrument :
MSVOA_D
ClientSampleId :





Tgt Ion:180 Resp: 3894
Ion Ratio Lower Upper
180 100
182 95.6 76.9 115.3
145 28.8 23.5 35.3




#71
Naphthalene
Concen: 1.190 ug/L
RT: 15.339 min Scan# 2338
Delta R.T. 0.006 min
Lab File: VD076436.D
Acq: 15 Jun 2023 13:20




Tgt Ion:128 Resp: 10768
Ion Ratio Lower Upper
128 100
127 14.5 10.8 16.2
129 7.7 8.6 12.8#





#72
 1,2,3-Trichlorobenzene
 Concen: 1.048 ug/L
 RT: 15.522 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VD076436.D
 Acq: 15 Jun 2023 13:20

Instrument :
 MSVOA_D
 ClientSampleId :



Tgt Ion:180 Resp: 4301
 Ion Ratio Lower Upper
 180 100
 182 84.0 74.3 111.5
 145 22.8 25.0 37.6#

