

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061523\
 Data File : VD076445.D
 Acq On : 15 Jun 2023 18:36
 Operator : KP/SY
 Sample : 03236-15
 Misc : 3.97G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 YBW23

Quant Time: Jun 16 02:00:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM061523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 16 01:50:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	309181	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	265855	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	124058	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	108922	14.466	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.880%
7) Chloroethane-d5	2.793	69	106995	17.574	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.280%
11) 1,1-Dichloroethene-d2	3.905	65	32516	16.544	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.160%
21) 2-Butanone-d5	6.987	46	62766	55.042	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.080%
24) Chloroform-d	7.569	84	169368	20.792	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.160%
26) 1,2-Dichloroethane-d4	8.228	65	87203	21.701	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.800%
32) Benzene-d6	8.199	84	337504	20.753	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	9.210	67	118881	22.901	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.600%
41) Toluene-d8	10.269	98	300274	19.941	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.760%
43) trans-1,3-Dichloroprop...	10.522	79	44251	20.495	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.960%
47) 2-Hexanone-d5	10.875	63	45491	56.454	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.900%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	94429	25.569	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.280%
66) 1,2-Dichlorobenzene-d4	13.816	152	103218	22.699	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.800%
Target Compounds						
					Qvalue	
16) Methylene chloride	4.793	84	9983	1.858	ug/L	96
34) Trichloroethene	9.022	95	2297	0.508	ug/L	# 81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061523\
Data File : VD076445.D
Acq On : 15 Jun 2023 18:36
Operator : KP/SY
Sample : 03236-15
Misc : 3.97G/10.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
YBW23

Quant Time: Jun 16 02:00:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM061523SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 16 01:50:34 2023
Response via : Initial Calibration









#16
Methylene chloride
Concen: 1.858 ug/L
RT: 4.793 min Scan# 544
Delta R.T. 0.006 min
Lab File: VD076445.D
Acq: 15 Jun 2023 18:36

Instrument :
MSVOA_D
ClientSampleId :
YBW23

Tgt Ion: 84 Resp: 9983
Ion Ratio Lower Upper
84 100
86 74.3 45.1 83.9
49 139.6 97.2 180.6



#34
Trichloroethene
Concen: 0.508 ug/L
RT: 9.022 min Scan# 1264
Delta R.T. -0.006 min
Lab File: VD076445.D
Acq: 15 Jun 2023 18:36

Tgt Ion: 95 Resp: 2297
Ion Ratio Lower Upper
95 100
97 90.8 43.1 80.1#
132 99.9 71.8 133.4
130 77.6 71.9 133.5

