

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061621\
 Data File : VD069465.D
 Acq On : 16 Jun 2021 14:25
 Operator : VA/SY
 Sample : M2709-07
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 0240-AMND-COMP-E

Quant Time: Jun 17 02:48:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 12:50:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	301683	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	584979	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	646052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	301029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	185068	54.376	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.760%	
35) Dibromofluoromethane	7.914	113	151912	34.671	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 69.340%	
50) Toluene-d8	10.338	98	787437	49.137	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.280%	
62) 4-Bromofluorobenzene	12.626	95	306405	59.590	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 119.180%	
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	77868	138.223	ug/l	100
17) Carbon Disulfide	4.409	76	115904	11.199	ug/l	98
25) 2-Butanone	7.214	43	14209	20.227	ug/l	93
39) Methylcyclohexane	9.356	83	21884	3.572	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061621\
Data File : VD069465.D
Acq On : 16 Jun 2021 14:25
Operator : VA/SY
Sample : M2709-07
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
0240-AMND-COMP-E

Quant Time: Jun 17 02:48:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 11 12:50:13 2021
Response via : Initial Calibration

