

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061623\
 Data File : VD076461.D
 Acq On : 16 Jun 2023 18:04
 Operator : KP/SY
 Sample : VIBLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Jun 19 04:56:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	165371	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	289281	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	257635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	123280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	81889	49.424	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.840%
35) Dibromofluoromethane	7.799	113	87670	51.015	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.020%
50) Toluene-d8	10.269	98	314725	48.290	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.580%
62) 4-Bromofluorobenzene	12.575	95	101902	48.176	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.360%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.799	84	24370	3.859	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061623\
Data File : VD076461.D
Acq On : 16 Jun 2023 18:04
Operator : KP/SY
Sample : VIBLK
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Jun 19 04:56:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

