

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061721\
 Data File : VD069476.D
 Acq On : 17 Jun 2021 13:27
 Operator : VA/SY
 Sample : M2735-01
 Misc : 6.83G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124027

Quant Time: Jun 18 01:25:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 12:50:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	54788	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	92382	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	101398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	46858	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	44789	72.463	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	144.920%
35) Dibromofluoromethane	7.914	113	40241	58.156	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.320%
50) Toluene-d8	10.337	98	116959	46.215	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.420%
62) 4-Bromofluorobenzene	12.626	95	39880	49.112	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061721\
Data File : VD069476.D
Acq On : 17 Jun 2021 13:27
Operator : VA/SY
Sample : M2735-01
Misc : 6.83G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124027

Quant Time: Jun 18 01:25:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 11 12:50:13 2021
Response via : Initial Calibration

