

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061721\
 Data File : VD069478.D
 Acq On : 17 Jun 2021 14:24
 Operator : VA/SY
 Sample : M2735-05
 Misc : 7.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124055

Quant Time: Jun 18 01:26:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 12:50:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	44196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	75681	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	84081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	35748	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	36820	73.847	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.700%
35) Dibromofluoromethane	7.920	113	31086	54.840	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.680%
50) Toluene-d8	10.338	98	95534	46.079	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.160%
62) 4-Bromofluorobenzene	12.626	95	33495	50.351	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061721\
Data File : VD069478.D
Acq On : 17 Jun 2021 14:24
Operator : VA/SY
Sample : M2735-05
Misc : 7.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124055

Quant Time: Jun 18 01:26:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 11 12:50:13 2021
Response via : Initial Calibration

