

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061721\
 Data File : VD069484.D
 Acq On : 17 Jun 2021 17:15
 Operator : VA/SY
 Sample : M2746-02
 Misc : 13.24G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NEW-DORP-2

Quant Time: Jun 18 01:27:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 12:50:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	249939	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	458029	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	479429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	181412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	164156	58.217	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.440%
35) Dibromofluoromethane	7.914	113	171254	49.919	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.840%
50) Toluene-d8	10.338	98	597489	47.618	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.240%
62) 4-Bromofluorobenzene	12.626	95	194748	48.372	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.740%
Target Compounds						
16) Acetone	4.156	43	565520	1211.673	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061721\
Data File : VD069484.D
Acq On : 17 Jun 2021 17:15
Operator : VA/SY
Sample : M2746-02
Misc : 13.24G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NEW-DORP-2

Quant Time: Jun 18 01:27:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 11 12:50:13 2021
Response via : Initial Calibration

