



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061721\
 Data File : VD069469.D
 Acq On : 17 Jun 2021 09:19
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061121S.M
 Title : SW846 8260
 Last Update : Fri Jun 11 12:50:13 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1811

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	36915	PASS
75	95	30	60	49.4	111587	PASS
95	95	100	100	100.0	226069	PASS
96	95	5	9	6.4	14524	PASS
173	174	0.00	2	1.6	2887	PASS
174	95	50	100	82.2	185728	PASS
175	174	5	9	7.2	13288	PASS
176	174	95	101	97.7	181525	PASS
177	176	5	9	6.7	12245	PASS