

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD061819\
 Data File : VD062775.D
 Acq On : 19 Jun 2019 5:34
 Operator : VA/AP
 Sample : K3157-09
 Misc : 6.42g/5ml/MSVOA D/SOIL
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-01-061719-E

Quant Time: Jun 19 07:28:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 00:02:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	531325	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.23	114	767050	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	465478	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	154834	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	387953	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
35) Dibromofluoromethane	6.93	113	349684	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
50) Toluene-d8	10.41	98	459917	27.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	54.06%	
62) 4-Bromofluorobenzene	13.55	95	181702	22.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	45.78%	
Target Compounds						
16) Acetone	3.23	43	109328	77.872	ug/l	97
18) Methyl Acetate	3.66	43	10504	4.553	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD061819\
Data File : VD062775.D
Acq On : 19 Jun 2019 5:34
Operator : VA/AP
Sample : K3157-09
Misc : 6.42g/5ml/MSVOA D/SOIL
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-01-061719-E

Quant Time: Jun 19 07:28:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060719S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 00:02:21 2019
Response via : Initial Calibration

