

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD061820\
 Data File : VD065815.D
 Acq On : 18 Jun 2020 18:22
 Operator : VA/SY
 Sample : L3024-13
 Misc : 5.88G/5.00ml/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-1

Quant Time: Jun 19 01:43:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060520S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 05 13:07:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.99	168	148336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	258415	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	220541	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	84405	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	85120	58.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.14%	
35) Dibromofluoromethane	7.91	113	16326	10.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	20.00%	
50) Toluene-d8	10.34	98	287325	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
62) 4-Bromofluorobenzene	12.64	95	78675	38.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.22%	
Target Compounds						
16) Acetone	4.16	43	6450	23.328	ug/l	92
20) Methylene Chloride	4.97	84	23603	12.761	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD061820\
Data File : VD065815.D
Acq On : 18 Jun 2020 18:22
Operator : VA/SY
Sample : L3024-13
Misc : 5.88G/5.00ml/MSVOA D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-1

Quant Time: Jun 19 01:43:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060520S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 05 13:07:30 2020
Response via : Initial Calibration

