


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD061919\
 Data File : VD062776.D
 Acq On : 19 Jun 2019 13:17
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
 Title : SW846 8260
 Last Update : Thu Jun 20 01:50:56 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.4	66968	PASS
75	95	30	60	46.7	139293	PASS
95	95	100	100	100.0	298517	PASS
96	95	5	9	6.0	17800	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	84.7	252778	PASS
175	174	5	9	6.7	16935	PASS
176	174	95	101	96.1	242922	PASS
177	176	5	9	6.8	16414	PASS