

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062018\
 Data File : VD058216.D
 Acq On : 20 Jun 2018 18:10
 Operator : VA/AP
 Sample : J3582-03
 Misc : 5.76g/5ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SSP-A-3-GRAB

Quant Time: Jun 21 04:19:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 20 05:52:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	667798	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.69	114	833577	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.69	117	660499	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.91	152	281455	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.02	65	343547	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
35) Dibromofluoromethane	5.55	113	367918	54.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.46%	
50) Toluene-d8	8.71	98	789658	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
62) 4-Bromofluorobenzene	11.93	95	297715	44.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	3.07	84	20315	4.06	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062018\
Data File : VD058216.D
Acq On : 20 Jun 2018 18:10
Operator : VA/AP
Sample : J3582-03
Misc : 5.76g/5ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SSP-A-3-GRAB

Quant Time: Jun 21 04:19:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 20 05:52:56 2018
Response via : Initial Calibration

