

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062019\
 Data File : VD062801.D
 Acq On : 20 Jun 2019 18:32
 Operator : VA/AP
 Sample : K3453-01RE
 Misc : 4.93g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FRAC-TANK-MUDRE

Quant Time: Jun 21 07:25:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 01:50:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.08	168	463522	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	8.25	114	612642	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.38	117	383868	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.54	152	116968	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.48	65	330394	65.57	ug/l	0.02
Spiked Amount 50.000			Recovery	=	131.14%	
35) Dibromofluoromethane	6.95	113	330025	63.89	ug/l	0.02
Spiked Amount 50.000			Recovery	=	127.78%	
50) Toluene-d8	10.44	98	509865	45.35	ug/l	0.03
Spiked Amount 50.000			Recovery	=	90.70%	
62) 4-Bromofluorobenzene	13.57	95	151483	29.63	ug/l	0.02
Spiked Amount 50.000			Recovery	=	59.26%	
Target Compounds						
16) Acetone	3.24	43	181297	175.349	ug/l	96
25) 2-Butanone	6.09	43	23484	24.921	ug/l	96
52) Toluene	10.53	92	10135	1.295	ug/l	94
64) Tetrachloroethene	11.29	164	6628	2.098	ug/l #	86
67) Ethyl Benzene	12.54	91	18249	1.427	ug/l	94
68) m/p-Xylenes	12.68	106	5767	1.295	ug/l	89
84) 1,2,4-Trimethylbenzene	14.25	105	7474	1.158	ug/l	94
95) Naphthalene	16.13	128	20912	6.174	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062019\
Data File : VD062801.D
Acq On : 20 Jun 2019 18:32
Operator : VA/AP
Sample : K3453-01RE
Misc : 4.93g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FRAC-TANK-MUDRE

Quant Time: Jun 21 07:25:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 01:50:56 2019
Response via : Initial Calibration

