

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062023\
 Data File : VD076487.D
 Acq On : 20 Jun 2023 12:46
 Operator : KP/SY
 Sample : 03289-16
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-12.0-14.0

Quant Time: Jun 21 06:45:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	191040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	346068	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	198104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	141052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	102098	53.341	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.680%	
35) Dibromofluoromethane	7.799	113	107612	52.344	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.680%	
50) Toluene-d8	10.269	98	213802	27.422	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 54.840%#	
62) 4-Bromofluorobenzene	12.575	95	121441	47.993	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 95.980%	
Target Compounds						
16) Acetone	4.017	43	6550	14.381	ug/l	Qvalue # 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062023\
Data File : VD076487.D
Acq On : 20 Jun 2023 12:46
Operator : KP/SY
Sample : 03289-16
Misc : 5.04G/5.0ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-12.0-14.0

Quant Time: Jun 21 06:45:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

