

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062023\
 Data File : VD076492.D
 Acq On : 20 Jun 2023 15:16
 Operator : KP/SY
 Sample : 03310-01
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-VPB-RW11-GW-968-970

Quant Time: Jun 20 15:30:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	143386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	242883	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	210167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	79223	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	71773	49.960	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.920%
35) Dibromofluoromethane	7.804	113	71085	49.266	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.540%
50) Toluene-d8	10.269	98	264800	48.391	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.780%
62) 4-Bromofluorobenzene	12.575	95	72194	40.651	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.300%
Target Compounds						
16) Acetone	4.028	43	6627	19.385	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062023\
Data File : VD076492.D
Acq On : 20 Jun 2023 15:16
Operator : KP/SY
Sample : 03310-01
Misc : 5.02G/5.0ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-VPB-RW11-GW-968-970

Quant Time: Jun 20 15:30:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

