

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062023\
 Data File : VD076496.D
 Acq On : 20 Jun 2023 17:17
 Operator : KP/SY
 Sample : 03289-19
 Misc : 5.09G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-07-0-2.0

Quant Time: Jun 21 06:48:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	159909	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	281569	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	250794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	98892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	93832	58.566	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.140%
35) Dibromofluoromethane	7.799	113	93454	55.870	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.740%
50) Toluene-d8	10.269	98	336247	53.005	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.000%
62) 4-Bromofluorobenzene	12.569	95	90835	44.120	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.240%
Target Compounds						
16) Acetone	4.023	43	8509	22.319	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062023\
Data File : VD076496.D
Acq On : 20 Jun 2023 17:17
Operator : KP/SY
Sample : 03289-19
Misc : 5.09G/5.0ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-07-0-2.0

Quant Time: Jun 21 06:48:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

