

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062023\
 Data File : VD076497.D
 Acq On : 20 Jun 2023 17:47
 Operator : KP/SY
 Sample : 03289-20
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-07-4.0-6.0

Quant Time: Jun 21 06:49:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	49446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	96528	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	101183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	50016	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	50287	101.506	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	203.020%	#
35) Dibromofluoromethane	7.804	113	38922	67.874	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	135.740%	
50) Toluene-d8	10.269	98	123823	56.936	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	113.880%	
62) 4-Bromofluorobenzene	12.575	95	44020	62.369	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	124.740%	
Target Compounds						
16) Acetone	4.016	43	13342	113.175	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062023\
Data File : VD076497.D
Acq On : 20 Jun 2023 17:47
Operator : KP/SY
Sample : 03289-20
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-07-4.0-6.0

Quant Time: Jun 21 06:49:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

