

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062118\
 Data File : VD058284.D
 Acq On : 22 Jun 2018 8:07
 Operator : VA/AP
 Sample : J3595-01
 Misc : 4.30g/5ml/MSVOA D/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SC-01-ABC

Quant Time: Jun 22 12:55:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 20 05:52:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	1660	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.70	114	3348	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.70	117	4469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.90	152	1936	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.02	65	2999	178.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	357.72%	
35) Dibromofluoromethane	5.56	113	927	34.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.04%	
50) Toluene-d8	8.71	98	4990	75.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	151.38%	
62) 4-Bromofluorobenzene	11.93	95	1987	73.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	147.56%	
Target Compounds						
6) Chloroethane	1.91	64	186	60.01	ug/l #	41
37) Ethyl Acetate	4.85	43	290	15.37	ug/l #	72
53) t-1,3-Dichloropropene	9.35	75	216	7.51	ug/l #	42
81) trans-1,4-Dichloro-2-buten	11.93	75	674	97.77	ug/l #	1
95) Naphthalene	14.52	128	188	2.91	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062118\
Data File : VD058284.D
Acq On : 22 Jun 2018 8:07
Operator : VA/AP
Sample : J3595-01
Misc : 4.30g/5ml/MSVOA D/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SC-01-ABC

Quant Time: Jun 22 12:55:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 20 05:52:56 2018
Response via : Initial Calibration

