

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062118\
 Data File : VD058288.D
 Acq On : 22 Jun 2018 10:26
 Operator : VA/AP
 Sample : J3582-07RE
 Misc : 6.02g/5ml/MSVOA D/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SSP-C-1-GRABRE

Quant Time: Jun 22 13:02:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 20 05:52:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	578247	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.70	114	760854	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.68	117	526273	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.91	152	194921	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.02	65	337333	57.76	ug/l	0.00
Spiked Amount			Recovery	=	115.52%	
35) Dibromofluoromethane	5.56	113	272039	43.93	ug/l	0.00
Spiked Amount			Recovery	=	87.86%	
50) Toluene-d8	8.71	98	723443	48.29	ug/l	0.00
Spiked Amount			Recovery	=	96.58%	
62) 4-Bromofluorobenzene	11.93	95	240557	39.31	ug/l	0.00
Spiked Amount			Recovery	=	78.62%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	3.08	84	17474	4.00	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062118\
Data File : VD058288.D
Acq On : 22 Jun 2018 10:26
Operator : VA/AP
Sample : J3582-07RE
Misc : 6.02g/5ml/MSVOA D/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SSP-C-1-GRABRE

Quant Time: Jun 22 13:02:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 20 05:52:56 2018
Response via : Initial Calibration

