

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062121\
 Data File : VD069508.D
 Acq On : 21 Jun 2021 19:18
 Operator : VA/SY
 Sample : M2793-03
 Misc : 6.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3

Quant Time: Jun 22 05:09:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	371942	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	734322	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	782100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	347053	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	285486	61.578	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.160%
35) Dibromofluoromethane	7.914	113	274824	55.365	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.720%
50) Toluene-d8	10.338	98	980734	52.673	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.340%
62) 4-Bromofluorobenzene	12.626	95	347053	56.697	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062121\
Data File : VD069508.D
Acq On : 21 Jun 2021 19:18
Operator : VA/SY
Sample : M2793-03
Misc : 6.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3

Quant Time: Jun 22 05:09:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

