

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
 Data File : VD076506.D
 Acq On : 21 Jun 2023 11:53
 Operator : KP/SY
 Sample : VD0621SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0621SBL01

Quant Time: Jun 22 01:55:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	113985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	206796	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	189129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	84638	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	54897	48.070	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.140%
35) Dibromofluoromethane	7.799	113	59257	48.235	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.460%
50) Toluene-d8	10.269	98	214954	46.137	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.280%
62) 4-Bromofluorobenzene	12.569	95	67282	44.497	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
Data File : VD076506.D
Acq On : 21 Jun 2023 11:53
Operator : KP/SY
Sample : VD0621SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0621SBL01

Quant Time: Jun 22 01:55:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

