

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
 Data File : VD076509.D
 Acq On : 21 Jun 2023 13:22
 Operator : KP/SY
 Sample : 03336-01
 Misc : 5.86G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-6

Quant Time: Jun 22 01:57:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	176040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	303635	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	267654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	103722	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	86955	49.301	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.600%
35) Dibromofluoromethane	7.804	113	93870	52.040	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.080%
50) Toluene-d8	10.269	98	350763	51.275	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.540%
62) 4-Bromofluorobenzene	12.575	95	99362	44.755	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
Data File : VD076509.D
Acq On : 21 Jun 2023 13:22
Operator : KP/SY
Sample : 03336-01
Misc : 5.86G/5.0ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-6

Quant Time: Jun 22 01:57:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

