

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
 Data File : VD076513.D
 Acq On : 21 Jun 2023 15:21
 Operator : KP/SY
 Sample : 03336-13
 Misc : 5.23G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-2

Quant Time: Jun 22 01:59:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	150140	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	265639	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	243102	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	102907	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	74932	49.813	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.620%
35) Dibromofluoromethane	7.805	113	81673	51.755	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.500%
50) Toluene-d8	10.269	98	312862	52.276	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.560%
62) 4-Bromofluorobenzene	12.575	95	91108	46.907	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
Data File : VD076513.D
Acq On : 21 Jun 2023 15:21
Operator : KP/SY
Sample : 03336-13
Misc : 5.23G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-2

Quant Time: Jun 22 01:59:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

