

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
 Data File : VD076523.D
 Acq On : 21 Jun 2023 20:22
 Operator : KP/SY
 Sample : 03355-09
 Misc : 5.07G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-05-6.5-7.0

Quant Time: Jun 22 02:03:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	48524	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	91186	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	82805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	30913	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	48958	100.702	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	201.400%#
35) Dibromofluoromethane	7.804	113	38920	71.847	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	143.700%
50) Toluene-d8	10.263	98	109024	53.068	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.140%
62) 4-Bromofluorobenzene	12.575	95	33024	49.530	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.060%
Target Compounds						
						Qvalue
16) Acetone	4.028	43	2438	21.074	ug/l #	73
20) Methylene Chloride	4.805	84	10444	9.872	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
Data File : VD076523.D
Acq On : 21 Jun 2023 20:22
Operator : KP/SY
Sample : 03355-09
Misc : 5.07G/5.0ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-05-6.5-7.0

Quant Time: Jun 22 02:03:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

