

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
 Data File : VD076524.D
 Acq On : 21 Jun 2023 20:52
 Operator : KP/SY
 Sample : 03355-11
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-6.5-7.0

Quant Time: Jun 22 02:03:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	59090	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	107838	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	100777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	36090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	53307	90.041	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 180.080%#	
35) Dibromofluoromethane	7.805	113	43606	68.067	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 136.140%	
50) Toluene-d8	10.269	98	123327	50.761	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.520%	
62) 4-Bromofluorobenzene	12.575	95	36931	46.837	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 93.680%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.793	84	11359	7.834	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062123\
Data File : VD076524.D
Acq On : 21 Jun 2023 20:52
Operator : KP/SY
Sample : 03355-11
Misc : 5.02G/5.0ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-6.5-7.0

Quant Time: Jun 22 02:03:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

